

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090524\
 Data File : VN083668.D
 Acq On : 05 Sep 2024 12:05
 Operator : JC\MD
 Sample : VN0905WBS01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0905WBS01

Quant Time: Sep 06 02:13:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083024W.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 31 00:48:10 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 09/06/2024
 Supervised By :Semsettin Yesilyurt 09/06/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	166869	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	286730	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	248061	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	122463	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	126286	54.150	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 108.300%			
35) Dibromofluoromethane	8.165	113	91750	53.165	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 106.320%			
50) Toluene-d8	10.565	98	334459	49.718	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 99.440%			
62) 4-Bromofluorobenzene	12.847	95	131132	52.673	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 105.340%			
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.124	85	40343	22.843	ug/l	96
3) Chloromethane	2.360	50	41834	20.335	ug/l	98
4) Vinyl Chloride	2.518	62	42238	20.515	ug/l	98
5) Bromomethane	2.901	94	23117	21.358	ug/l	100
6) Chloroethane	3.071	64	29121	21.644	ug/l	98
7) Trichlorofluoromethane	3.465	101	73901	21.796	ug/l	98
8) Diethyl Ether	3.954	74	25194	20.785	ug/l	98
9) 1,1,2-Trichlorotrifluo...	4.359	101	38988	21.203	ug/l	99
10) Methyl Iodide	4.571	142	51666	20.771	ug/l #	95
11) Tert butyl alcohol	5.530	59	34492	94.856	ug/l	99
12) 1,1-Dichloroethene	4.324	96	37544	20.537	ug/l	91
13) Acrolein	4.171	56	26243	76.284	ug/l	99
14) Allyl chloride	5.012	41	63126	19.368	ug/l	97
15) Acrylonitrile	5.718	53	89937	100.292	ug/l	98
16) Acetone	4.430	43	88785	95.466	ug/l	99
17) Carbon Disulfide	4.701	76	97071	18.222	ug/l #	94
18) Methyl Acetate	5.018	43	43760	17.693	ug/l	99
19) Methyl tert-butyl Ether	5.795	73	131516	20.282	ug/l	98
20) Methylene Chloride	5.265	84	42284	20.589	ug/l	91
21) trans-1,2-Dichloroethene	5.777	96	37759	19.829	ug/l	92
22) Diisopropyl ether	6.665	45	130582	20.318	ug/l	98
23) Vinyl Acetate	6.600	43	487914	102.443	ug/l	99
24) 1,1-Dichloroethane	6.559	63	75358	20.561	ug/l	98
25) 2-Butanone	7.483	43	122545	96.778	ug/l	98
26) 2,2-Dichloropropane	7.483	77	70975	20.800	ug/l	99
27) cis-1,2-Dichloroethene	7.483	96	46892	20.637	ug/l	96
28) Bromochloromethane	7.806	49	36215	21.160	ug/l	99
29) Tetrahydrofuran	7.841	42	77703	98.931	ug/l	99
30) Chloroform	7.965	83	77311	20.022	ug/l	98
31) Cyclohexane	8.253	56	67214	18.632	ug/l	96
32) 1,1,1-Trichloroethane	8.165	97	73190	20.961	ug/l	99
36) 1,1-Dichloropropene	8.365	75	53575	20.353	ug/l	98
37) Ethyl Acetate	7.553	43	53425	20.393	ug/l	98
38) Carbon Tetrachloride	8.359	117	63737	21.222	ug/l	94
39) Methylcyclohexane	9.594	83	65584	20.763	ug/l	97
40) Benzene	8.600	78	165831	20.479	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090524\
 Data File : VN083668.D
 Acq On : 05 Sep 2024 12:05
 Operator : JC\MD
 Sample : VN0905WBS01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0905WBS01

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 09/06/2024
 Supervised By :Semsettin Yesilyurt 09/06/2024

Quant Time: Sep 06 02:13:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083024W.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 31 00:48:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.771	41	28779	19.251	ug/l	97
42) 1,2-Dichloroethane	8.665	62	62217	20.728	ug/l	99
43) Isopropyl Acetate	8.688	43	90096	16.343	ug/l	97
44) Trichloroethene	9.353	130	37894	19.453	ug/l	100
45) 1,2-Dichloropropane	9.618	63	39642	21.015	ug/l	99
46) Dibromomethane	9.706	93	28922	20.783	ug/l	99
47) Bromodichloromethane	9.888	83	59372	20.042	ug/l	97
48) Methyl methacrylate	9.677	41	44649	20.779	ug/l	98
49) 1,4-Dioxane	9.694	88	14534	412.793	ug/l #	98
51) 4-Methyl-2-Pentanone	10.447	43	248410	101.993	ug/l	98
52) Toluene	10.630	92	103542	20.445	ug/l	98
53) t-1,3-Dichloropropene	10.835	75	61139	19.936	ug/l	98
54) cis-1,3-Dichloropropene	10.312	75	65791	20.439	ug/l	97
55) 1,1,2-Trichloroethane	11.012	97	36897	20.412	ug/l	97
56) Ethyl methacrylate	10.871	69	61803	20.668	ug/l	98
57) 1,3-Dichloropropane	11.159	76	66495	20.779	ug/l	98
58) 2-Chloroethyl Vinyl ether	10.159	63	117307	90.282	ug/l	98
59) 2-Hexanone	11.194	43	184982	101.203	ug/l	98
60) Dibromochloromethane	11.359	129	43172	20.612	ug/l	99
61) 1,2-Dibromoethane	11.471	107	37990	20.571	ug/l	97
64) Tetrachloroethene	11.106	164	31588	19.373	ug/l	95
65) Chlorobenzene	11.888	112	108152	19.441	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.959	131	37986	20.224	ug/l	99
67) Ethyl Benzene	11.965	91	196379	19.776	ug/l	99
68) m/p-Xylenes	12.071	106	146362	39.125	ug/l	100
69) o-Xylene	12.394	106	70552	19.513	ug/l	99
70) Styrene	12.412	104	115567	19.010	ug/l	99
71) Bromoform	12.576	173	26457	19.776	ug/l #	96
73) Isopropylbenzene	12.694	105	187473	18.914	ug/l	100
74) N-amyl acetate	12.494	43	79210	18.412	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.935	83	53865	19.142	ug/l	97
76) 1,2,3-Trichloropropane	12.994	75	39236m	15.869	ug/l	
77) Bromobenzene	12.976	156	42218	18.155	ug/l	98
78) n-propylbenzene	13.035	91	210253	18.347	ug/l	99
79) 2-Chlorotoluene	13.124	91	132660	17.864	ug/l	98
80) 1,3,5-Trimethylbenzene	13.171	105	157248	18.832	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.735	75	17467	16.662	ug/l	94
82) 4-Chlorotoluene	13.224	91	131027	17.368	ug/l	99
83) tert-Butylbenzene	13.435	119	141609	19.075	ug/l	98
84) 1,2,4-Trimethylbenzene	13.482	105	156571	18.629	ug/l	99
85) sec-Butylbenzene	13.612	105	181338	18.378	ug/l	100
86) p-Isopropyltoluene	13.729	119	153119	18.579	ug/l	99
87) 1,3-Dichlorobenzene	13.735	146	76397	16.924	ug/l	98
88) 1,4-Dichlorobenzene	13.812	146	77693	16.750	ug/l	99
89) n-Butylbenzene	14.053	91	126411	17.438	ug/l	100
90) Hexachloroethane	14.329	117	28290	17.635	ug/l	98
91) 1,2-Dichlorobenzene	14.106	146	78118	17.970	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.718	75	10501	17.207	ug/l	98
93) 1,2,4-Trichlorobenzene	15.394	180	37302	15.132	ug/l	97
94) Hexachlorobutadiene	15.500	225	17227	17.450	ug/l	96
95) Naphthalene	15.635	128	123142	15.548	ug/l	100
96) 1,2,3-Trichlorobenzene	15.841	180	35998	14.618	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090524\
Data File : VN083668.D
Acq On : 05 Sep 2024 12:05
Operator : JC\MD
Sample : VN0905WBS01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024

Quant Time: Sep 06 02:13:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083024W.M
Quant Title : SW846 8260
QLast Update : Sat Aug 31 00:48:10 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090524\
Data File : VN083668.D
Acq On : 05 Sep 2024 12:05
Operator : JC\MD
Sample : VN0905WBS01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024

Quant Time: Sep 06 02:13:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083024W.M
Quant Title : SW846 8260
QLast Update : Sat Aug 31 00:48:10 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.218 min Scan# 1065
Delta R.T. -0.006 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05

Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01

Tgt Ion:168 Resp: 166869
Ion Ratio Lower Upper
168 100
99 68.6 54.2 81.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024



#2
Dichlorodifluoromethane
Concen: 22.843 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05


Tgt Ion: 85 Resp: 40343
Ion Ratio Lower Upper
85 100
87 33.6 15.7 47.0





#3
Chloromethane
Concen: 20.335 ug/l
RT: 2.360 min Scan# 69
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01



Tgt Ion: 50 Resp: 41834
Ion Ratio Lower Upper
50 100
52 34.9 27.2 40.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024



#4
Vinyl Chloride
Concen: 20.515 ug/l
RT: 2.518 min Scan# 96
Delta R.T. 0.006 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05

Tgt Ion: 62 Resp: 42238
Ion Ratio Lower Upper
62 100
64 32.0 24.7 37.1









#5
Bromomethane
Concen: 21.358 ug/l
RT: 2.901 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05

Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01

Tgt Ion: 94 Resp: 2311
Ion Ratio Lower Upper
94 100
96 94.6 76.0 114.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024



#6
Chloroethane
Concen: 21.644 ug/l
RT: 3.071 min Scan# 190
Delta R.T. -0.006 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05

Tgt Ion: 64 Resp: 29121
Ion Ratio Lower Upper
64 100
66 31.0 23.9 35.9









#7
Trichlorofluoromethane
Concen: 21.796 ug/l
RT: 3.465 min Scan# 21
Delta R.T. -0.006 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05

Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01

Tgt Ion:101 Resp: 7390
Ion Ratio Lower Upper
101 100
103 62.5 51.4 77.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024



#8
Diethyl Ether
Concen: 20.785 ug/l
RT: 3.954 min Scan# 340
Delta R.T. -0.006 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05

Tgt Ion: 74 Resp: 25194
Ion Ratio Lower Upper
74 100
45 91.5 46.7 140.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 21.203 ug/l
RT: 4.359 min Scan# 408
Delta R.T. 0.006 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05

Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01

Tgt Ion:101 Resp: 3898
Ion Ratio Lower Upper
101 100
85 46.6 36.2 54.4
151 71.4 57.7 86.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024



#10
Methyl Iodide
Concen: 20.771 ug/l
RT: 4.571 min Scan# 445
Delta R.T. -0.006 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05


Tgt Ion:142 Resp: 51666
Ion Ratio Lower Upper
142 100
127 49.3 37.5 56.3
141 17.6 11.0 16.4#





#11
Tert butyl alcohol
Concen: 94.856 ug/l
RT: 5.530 min Scan# 608
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01




Tgt Ion: 59 Resp: 3449
Ion Ratio Lower Upper
59 100
57 10.5 8.7 13.1

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024




#12
1,1-Dichloroethene
Concen: 20.537 ug/l
RT: 4.324 min Scan# 403
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Tgt Ion: 96 Resp: 37544
Ion Ratio Lower Upper
96 100
61 168.2 143.4 215.0
98 56.2 52.7 79.1





#13
Acrolein
Concen: 76.284 ug/l
RT: 4.171 min Scan# 31
Delta R.T. -0.006 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01



Tgt Ion: 56 Resp: 2624
Ion Ratio Lower Upper
56 100
55 71.7 58.2 87.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024



#14
Allyl chloride
Concen: 19.368 ug/l
RT: 5.012 min Scan# 520
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05


Tgt Ion: 41 Resp: 63126
Ion Ratio Lower Upper
41 100
39 77.4 59.3 88.9
76 35.8 27.7 41.5





#15
Acrylonitrile
Concen: 100.292 ug/l
RT: 5.718 min Scan# 640
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01



Tgt Ion: 53 Resp: 8993
Ion Ratio Lower Upper
53 100
52 81.1 66.1 99.1
51 35.6 29.6 44.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024



#16
Acetone
Concen: 95.466 ug/l
RT: 4.430 min Scan# 421
Delta R.T. 0.006 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05


Tgt Ion: 43 Resp: 88785
Ion Ratio Lower Upper
43 100
58 28.9 23.7 35.5





#17
Carbon Disulfide
Concen: 18.222 ug/l
RT: 4.701 min Scan# 467
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01



Tgt Ion: 76 Resp: 9707
Ion Ratio Lower Upper
76 100
78 10.9 7.1 10.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024



#18
Methyl Acetate
Concen: 17.693 ug/l
RT: 5.018 min Scan# 521
Delta R.T. -0.006 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05


Tgt Ion: 43 Resp: 43760
Ion Ratio Lower Upper
43 100
74 24.1 19.0 28.4





#19
Methyl tert-butyl Ether
Concen: 20.282 ug/l
RT: 5.795 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01



Tgt Ion: 73 Resp: 131510
Ion Ratio Lower Upper
73 100
57 22.1 17.0 25.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024



#20
Methylene Chloride
Concen: 20.589 ug/l
RT: 5.265 min Scan# 563
Delta R.T. -0.006 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05


Tgt Ion: 84 Resp: 42284
Ion Ratio Lower Upper
84 100
49 131.3 93.4 140.0
51 34.0 29.2 43.8
86 59.7 46.9 70.3





#21
trans-1,2-Dichloroethene
Concen: 19.829 ug/l
RT: 5.777 min Scan# 61
Delta R.T. -0.006 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01



Tgt Ion: 96 Resp: 3775
Ion Ratio Lower Upper
96 100
61 145.2 106.0 159.0
98 66.4 52.2 78.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024



#22
Diisopropyl ether
Concen: 20.318 ug/l
RT: 6.665 min Scan# 801
Delta R.T. -0.006 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05

Tgt Ion: 45 Resp: 130582
Ion Ratio Lower Upper
45 100
43 55.5 43.4 65.0
87 27.2 22.6 33.8
59 12.3 9.1 13.7









#23
 Vinyl Acetate
 Concen: 102.443 ug/l
 RT: 6.600 min Scan# 790
 Delta R.T. 0.000 min
 Lab File: VN083668.D
 Acq: 05 Sep 2024 12:05

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0905WBS01

Tgt Ion: 43 Resp: 487914
 Ion Ratio Lower Upper
 43 100
 86 10.6 8.7 13.1




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 09/06/2024
 Supervised By :Semsettin Yesilyurt 09/06/2024



#24
 1,1-Dichloroethane
 Concen: 20.561 ug/l
 RT: 6.559 min Scan# 783
 Delta R.T. -0.006 min
 Lab File: VN083668.D
 Acq: 05 Sep 2024 12:05


Tgt Ion: 63 Resp: 75358
 Ion Ratio Lower Upper
 63 100
 98 6.2 3.5 10.4
 100 6.0 2.5 7.5





#25
2-Butanone
Concen: 96.778 ug/l
RT: 7.483 min Scan# 940
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01



Tgt Ion: 43 Resp: 122549
Ion Ratio Lower Upper
43 100
72 24.9 20.8 31.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024



#26
2,2-Dichloropropane
Concen: 20.800 ug/l
RT: 7.483 min Scan# 940
Delta R.T. -0.006 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05


Tgt Ion: 77 Resp: 70975
Ion Ratio Lower Upper
77 100
97 20.9 10.2 30.4





#27
cis-1,2-Dichloroethene
Concen: 20.637 ug/l
RT: 7.483 min Scan# 940
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01



Tgt Ion: 96 Resp: 46892
Ion Ratio Lower Upper
96 100
61 147.2 0.0 306.2
98 63.1 0.0 128.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024



#28
Bromochloromethane
Concen: 21.160 ug/l
RT: 7.806 min Scan# 995
Delta R.T. -0.006 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05


Tgt Ion: 49 Resp: 36215
Ion Ratio Lower Upper
49 100
129 1.8 0.0 3.4
130 75.8 60.2 90.2





#29
Tetrahydrofuran
Concen: 98.931 ug/l
RT: 7.841 min Scan# 1000
Delta R.T. 0.006 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01



Tgt Ion: 42 Resp: 7770
Ion Ratio Lower Upper
42 100
72 45.4 35.9 53.9
71 41.6 33.9 50.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024



#30
Chloroform
Concen: 20.022 ug/l
RT: 7.965 min Scan# 1022
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05


Tgt Ion: 83 Resp: 77311
Ion Ratio Lower Upper
83 100
85 64.7 52.8 79.2





#31
Cyclohexane
Concen: 18.632 ug/l
RT: 8.253 min Scan# 1071
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01




Tgt Ion: 56 Resp: 67214
Ion Ratio Lower Upper
56 100
69 34.7 26.9 40.3
84 90.8 69.2 103.8

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024




#32
1,1,1-Trichloroethane
Concen: 20.961 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Tgt Ion: 97 Resp: 73190
Ion Ratio Lower Upper
97 100
99 66.1 52.3 78.5
61 49.0 38.2 57.2





#33
1,2-Dichloroethane-d4
Concen: 54.150 ug/l
RT: 8.577 min Scan# 1126
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01




Tgt Ion: 65 Resp: 126280
Ion Ratio Lower Upper
65 100
67 49.6 0.0 100.8

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.100 min Scan# 1215
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Tgt Ion:114 Resp: 286730
Ion Ratio Lower Upper
114 100
63 21.7 0.0 43.4
88 17.3 0.0 32.6





#35
Dibromofluoromethane
Concen: 53.165 ug/l
RT: 8.165 min Scan# 1090
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01




Tgt Ion: 113 Resp: 91750
Ion Ratio Lower Upper
113 100
111 104.9 82.6 124.0
192 17.3 14.3 21.5

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024




#36
1,1-Dichloropropene
Concen: 20.353 ug/l
RT: 8.365 min Scan# 1090
Delta R.T. -0.006 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05



Tgt Ion: 75 Resp: 53575
Ion Ratio Lower Upper
75 100
110 34.3 16.7 50.1
77 30.6 25.2 37.8









#37
Ethyl Acetate
Concen: 20.393 ug/l
RT: 7.553 min Scan# 91
Delta R.T. -0.006 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05

Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01

Tgt Ion: 43 Resp: 5342
Ion Ratio Lower Upper
43 100
61 13.3 11.1 16.7
70 10.1 8.8 13.2




Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024



#38
Carbon Tetrachloride
Concen: 21.222 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05

Tgt Ion:117 Resp: 63737
Ion Ratio Lower Upper
117 100
119 102.3 76.1 114.1
121 29.3 24.1 36.1





#39

Methylcyclohexane

Concen: 20.763 ug/l

RT: 9.594 min Scan# 1130

Delta R.T. -0.006 min

Lab File: VN083668.D

Acq: 05 Sep 2024 12:05

Instrument :

MSVOA_N

ClientSampleId :

VN0905WBS01





Tgt Ion: 83 Resp: 65584

Ion	Ratio	Lower	Upper
83	100		
55	73.8	62.2	93.4
98	45.9	37.6	56.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024



#40

Benzene

Concen: 20.479 ug/l

RT: 8.600 min Scan# 1130




Delta R.T. -0.006 min

Lab File: VN083668.D

Acq: 05 Sep 2024 12:05

Tgt Ion: 78 Resp: 165831


Ion	Ratio	Lower	Upper
78	100		
77	22.6	19.0	28.6





#41
Methacrylonitrile
Concen: 19.251 ug/l
RT: 7.771 min Scan# 91
Delta R.T. -0.006 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01



Tgt Ion:	41	Resp:	28779
Ion	Ratio	Lower	Upper
41	100		
39	55.6	45.4	68.2
67	72.3	55.9	83.9
52	32.7	24.1	36.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024



#42
1,2-Dichloroethane
Concen: 20.728 ug/l
RT: 8.665 min Scan# 1141
Delta R.T. -0.006 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05

Tgt Ion:	62	Resp:	62217
Ion	Ratio	Lower	Upper
62	100		
98	8.0	0.0	16.6





#43

Isopropyl Acetate

Concen: 16.343 ug/l

RT: 8.688 min Scan# 1145

Delta R.T. 0.000 min

Lab File: VN083668.D

Acq: 05 Sep 2024 12:05

Instrument :

MSVOA_N

ClientSampleId :





VN0905WBS01

Tgt Ion: 43	Resp: 90090
Ion Ratio	Lower Upper
43 100	
61 25.7	19.4 29.0
87 13.4	9.8 14.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024



#44

Trichloroethene

Concen: 19.453 ug/l




RT: 9.353 min Scan# 1258

Delta R.T. 0.006 min

Lab File: VN083668.D

Acq: 05 Sep 2024 12:05

Tgt Ion:130	Resp: 37894
Ion Ratio	Lower Upper
130 100	
95 100.3	0.0 200.4





#45

1,2-Dichloropropane

Concen: 21.015 ug/l

RT: 9.618 min Scan# 1303

Delta R.T. 0.000 min

Lab File: VN083668.D

Acq: 05 Sep 2024 12:05

Instrument :

MSVOA_N

ClientSampleId :

VN0905WBS01

Tgt Ion: 63 Resp: 3964

Ion Ratio Lower Upper

63 100





65 32.9 25.9 38.9

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/06/2024

Supervised By :Semsettin Yesilyurt 09/06/2024



#46

Dibromomethane

Concen: 20.783 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. 0.000 min

Lab File: VN083668.D

Acq: 05 Sep 2024 12:05




Tgt Ion: 93 Resp: 28922


Ion Ratio Lower Upper

93 100

95 80.9 65.3 97.9

174 88.4 71.0 106.4









#47
Bromodichloromethane
Concen: 20.042 ug/l
RT: 9.888 min Scan# 1349
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05

Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01

Tgt Ion: 83 Resp: 59372
Ion Ratio Lower Upper
83 100
85 62.9 51.8 77.8
127 7.0 6.6 9.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024



#48
Methyl methacrylate
Concen: 20.779 ug/l
RT: 9.677 min Scan# 1313
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05

Tgt Ion: 41 Resp: 44649
Ion Ratio Lower Upper
41 100
69 78.5 64.3 96.5
39 60.4 48.9 73.3





#49

1,4-Dioxane

Concen: 412.793 ug/l

RT: 9.694 min Scan# 1316

Delta R.T. 0.000 min

Lab File: VN083668.D

Acq: 05 Sep 2024 12:05

Instrument :

MSVOA_N

ClientSampleId :

VN0905WBS01

Tgt Ion: 88 Resp: 14534

Ion Ratio Lower Upper

88 100

43 32.1 0.0 0.0





58 73.1 57.1 85.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/06/2024

Supervised By :Semsettin Yesilyurt 09/06/2024



#50

Toluene-d8

Concen: 49.718 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.000 min

Lab File: VN083668.D




Acq: 05 Sep 2024 12:05


Tgt Ion: 98 Resp: 334459

Ion Ratio Lower Upper

98 100

100 65.4 52.0 78.0









#51
4-Methyl-2-Pentanone
Concen: 101.993 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05

Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01

Tgt Ion: 43 Resp: 248410
Ion Ratio Lower Upper
43 100
58 38.2 31.6 47.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024



#52
Toluene
Concen: 20.445 ug/l
RT: 10.630 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05


Tgt Ion: 92 Resp: 103542
Ion Ratio Lower Upper
92 100
91 171.5 134.7 202.1





#53
t-1,3-Dichloropropene
Concen: 19.936 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01



Tgt Ion: 75 Resp: 61139
Ion Ratio Lower Upper
75 100
77 30.9 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024



#54
cis-1,3-Dichloropropene
Concen: 20.439 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05


Tgt Ion: 75 Resp: 65791
Ion Ratio Lower Upper
75 100
77 31.5 25.0 37.6
39 57.0 42.6 64.0





#55
1,1,2-Trichloroethane
Concen: 20.412 ug/l
RT: 11.012 min Scan# 1541
Delta R.T. -0.006 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01



Tgt Ion	Ratio	Lower	Upper
97	100		
83	78.7	67.5	101.3
85	55.8	44.2	66.2
99	63.4	50.1	75.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024



#56
Ethyl methacrylate
Concen: 20.668 ug/l
RT: 10.871 min Scan# 1516
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05

Tgt Ion	Ratio	Lower	Upper
69	100		
41	68.0	53.9	80.9
39	43.7	32.6	48.8





#57

1,3-Dichloropropane

Concen: 20.779 ug/l

RT: 11.159 min Scan# 1565

Delta R.T. 0.000 min

Lab File: VN083668.D

Acq: 05 Sep 2024 12:05

Instrument :

MSVOA_N

ClientSampleId :

VN0905WBS01

Tgt Ion: 76 Resp: 6649

Ion Ratio Lower Upper

76 100





78 33.4 25.8 38.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/06/2024

Supervised By :Semsettin Yesilyurt 09/06/2024



#58

2-Chloroethyl Vinyl ether

Concen: 90.282 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. 0.000 min

Lab File: VN083668.D




Acq: 05 Sep 2024 12:05


Tgt Ion: 63 Resp: 117307

Ion Ratio Lower Upper

63 100


106 28.2 23.3 34.9





#59
2-Hexanone
Concen: 101.203 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01



Tgt Ion: 43 Resp: 18498
Ion Ratio Lower Upper
43 100
58 52.7 27.2 81.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024



#60
Dibromochloromethane
Concen: 20.612 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05


Tgt Ion:129 Resp: 43172
Ion Ratio Lower Upper
129 100
127 77.3 39.0 116.9





#61
1,2-Dibromoethane
Concen: 20.571 ug/l
RT: 11.471 min Scan# 1618
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01




Tgt Ion: 107 Resp: 37990
Ion Ratio Lower Upper
107 100
109 92.1 75.6 113.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024




#62
4-Bromofluorobenzene
Concen: 52.673 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05



Tgt Ion: 95 Resp: 131132
Ion Ratio Lower Upper
95 100
174 70.3 0.0 139.8
176 69.1 0.0 131.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN083668.D

Acq: 05 Sep 2024 12:05

Instrument :

MSVOA_N

ClientSampleId :

VN0905WBS01

Tgt Ion:117 Resp: 24806

Ion Ratio Lower Upper

117 100

82 57.7 47.4 71.0





119 32.3 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/06/2024

Supervised By :Semsettin Yesilyurt 09/06/2024



#64

Tetrachloroethene

Concen: 19.373 ug/l

RT: 11.106 min Scan# 1556

Delta R.T. 0.006 min

Lab File: VN083668.D

Acq: 05 Sep 2024 12:05

Tgt Ion:164 Resp: 31588




Ion Ratio Lower Upper


164 100

166 122.9 103.7 155.5

129 94.0 79.9 119.9


131 94.1 78.4 117.6





#65
Chlorobenzene
Concen: 19.441 ug/l
RT: 11.888 min Scan# 110
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01



Tgt Ion:112 Resp: 108152
Ion Ratio Lower Upper
112 100
114 31.3 25.7 38.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024



#66
1,1,1,2-Tetrachloroethane
Concen: 20.224 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05


Tgt Ion:131 Resp: 37986
Ion Ratio Lower Upper
131 100
133 96.1 47.5 142.6
119 63.2 31.6 94.8





#67
Ethyl Benzene
Concen: 19.776 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01




Tgt Ion: 91 Resp: 196379
Ion Ratio Lower Upper
91 100
106 29.9 24.3 36.5

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024




#68
m/p-Xylenes
Concen: 39.125 ug/l
RT: 12.071 min Scan# 1720
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05



Tgt Ion:106 Resp: 146362
Ion Ratio Lower Upper
106 100
91 212.0 169.2 253.8









#69
o-Xylene
Concen: 19.513 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05

Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01

Tgt Ion:106 Resp: 70552
Ion Ratio Lower Upper
106 100
91 227.6 112.9 338.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024



#70
Styrene
Concen: 19.010 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05


Tgt Ion:104 Resp: 115567
Ion Ratio Lower Upper
104 100
78 55.4 43.3 64.9
103 54.6 43.4 65.2





#71
Bromoform
Concen: 19.776 ug/l
RT: 12.576 min Scan# 1806
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01




Tgt Ion:173 Resp: 2645
Ion Ratio Lower Upper
173 100
175 50.7 24.1 72.2
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Tgt Ion:152 Resp: 122463
Ion Ratio Lower Upper
152 100
115 63.8 32.2 96.6
150 165.0 0.0 348.6





#73
Isopropylbenzene
Concen: 18.914 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01




Tgt Ion:105 Resp: 18747
Ion Ratio Lower Upper
105 100
120 25.2 12.6 37.8

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024




#74
N-ethyl acetate
Concen: 18.412 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Tgt Ion: 43 Resp: 79210
Ion Ratio Lower Upper
43 100
70 42.8 34.5 51.7
55 24.7 18.6 28.0
61 23.7 19.2 28.8





#75
1,1,2,2-Tetrachloroethane
Concen: 19.142 ug/l
RT: 12.935 min Scan# 1867
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01



Tgt Ion: 83 Resp: 5386
Ion Ratio Lower Upper
83 100
131 9.5 5.2 15.6
85 61.5 31.9 95.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024



#76
1,2,3-Trichloropropane
Concen: 15.869 ug/l m
RT: 12.994 min Scan# 1877
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05


Tgt Ion: 75 Resp: 39236
Ion Ratio Lower Upper
75 100
77 242.9 95.2 285.6





#77
Bromobenzene
Concen: 18.155 ug/l
RT: 12.976 min Scan# 1875
Delta R.T. -0.006 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01



Tgt Ion:156 Resp: 42218
Ion Ratio Lower Upper
156 100
77 225.7 114.3 343.1
158 95.9 48.9 146.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024



#78
n-propylbenzene
Concen: 18.347 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05

Tgt Ion: 91 Resp: 210253
Ion Ratio Lower Upper
91 100
120 21.4 11.0 33.0









#79
2-Chlorotoluene
Concen: 17.864 ug/l
RT: 13.124 min Scan# 1899
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05

Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01

Tgt Ion: 91 Resp: 132660
Ion Ratio Lower Upper
91 100
126 32.1 15.4 46.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024



#80
1,3,5-Trimethylbenzene
Concen: 18.832 ug/l
RT: 13.171 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05


Tgt Ion:105 Resp: 157248
Ion Ratio Lower Upper
105 100
120 46.2 23.4 70.3





#81
trans-1,4-Dichloro-2-butene
Concen: 16.662 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01



Tgt Ion: 75 Resp: 1746
Ion Ratio Lower Upper
75 100
53 108.0 81.2 121.8
89 48.3 35.7 53.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024



#82
4-Chlorotoluene
Concen: 17.368 ug/l
RT: 13.224 min Scan# 1916
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05


Tgt Ion: 91 Resp: 131027
Ion Ratio Lower Upper
91 100
126 31.5 15.4 46.2





#83
tert-Butylbenzene
Concen: 19.075 ug/l
RT: 13.435 min Scan# 1952
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01




Tgt Ion:119 Resp: 141609
Ion Ratio Lower Upper
119 100
91 69.7 34.5 103.6
134 23.5 12.8 38.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024




#84
1,2,4-Trimethylbenzene
Concen: 18.629 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Tgt Ion:105 Resp: 156571
Ion Ratio Lower Upper
105 100
120 43.0 21.8 65.4





#85
 sec-Butylbenzene
 Concen: 18.378 ug/l
 RT: 13.612 min Scan# 1983
 Delta R.T. -0.006 min
 Lab File: VN083668.D
 Acq: 05 Sep 2024 12:05




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0905WBS01




Tgt Ion:105 Resp: 181338
 Ion Ratio Lower Upper
 105 100
 134 19.5 9.7 29.1

Manual Integrations
 APPROVED



Reviewed By :Mahesh Dadoda 09/06/2024
 Supervised By :Semsettin Yesilyurt 09/06/2024




#86
 p-Isopropyltoluene
 Concen: 18.579 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. 0.000 min
 Lab File: VN083668.D
 Acq: 05 Sep 2024 12:05




Tgt Ion:119 Resp: 153119
 Ion Ratio Lower Upper
 119 100
 134 24.5 12.8 38.3
 91 26.2 12.9 38.7





#87
1,3-Dichlorobenzene
Concen: 16.924 ug/l
RT: 13.735 min Scan# 2003
Delta R.T. 0.006 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01



Tgt Ion:146 Resp: 7639
Ion Ratio Lower Upper
146 100
111 44.8 21.7 65.1
148 64.7 31.7 95.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024



#88
1,4-Dichlorobenzene
Concen: 16.750 ug/l
RT: 13.812 min Scan# 2016
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05

Tgt Ion:146 Resp: 77693
Ion Ratio Lower Upper
146 100
111 43.8 21.3 64.0
148 64.7 32.0 96.2





#89

n-Butylbenzene

Concen: 17.438 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. 0.000 min

Lab File: VN083668.D

Acq: 05 Sep 2024 12:05

Instrument :

MSVOA_N

ClientSampleId :

VN0905WBS01





Tgt Ion: 91 Resp: 12641

Ion	Ratio	Lower	Upper
91	100		
92	53.7	26.7	80.1
134	24.4	12.2	36.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024



#90

Hexachloroethane

Concen: 17.635 ug/l

RT: 14.329 min Scan# 2104




Delta R.T. 0.000 min

Lab File: VN083668.D

Acq: 05 Sep 2024 12:05

Tgt Ion:117 Resp: 28290


Ion	Ratio	Lower	Upper
117	100		
201	67.2	32.8	98.3





#91
1,2-Dichlorobenzene
Concen: 17.970 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01




Tgt Ion:146 Resp: 78118
Ion Ratio Lower Upper
146 100
111 44.0 22.8 68.3
148 62.7 32.1 96.5

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024




#92
1,2-Dibromo-3-Chloropropane
Concen: 17.207 ug/l
RT: 14.718 min Scan# 2170
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Tgt Ion: 75 Resp: 10501
Ion Ratio Lower Upper
75 100
155 75.2 35.9 107.7
157 94.4 47.5 142.5





#93
1,2,4-Trichlorobenzene
Concen: 15.132 ug/l
RT: 15.394 min Scan# 2285
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01




Tgt Ion:180 Resp: 37302
Ion Ratio Lower Upper
180 100
182 92.9 48.1 144.3
145 34.3 16.9 50.7

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024




#94
Hexachlorobutadiene
Concen: 17.450 ug/l
RT: 15.500 min Scan# 2303
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Tgt Ion:225 Resp: 17227
Ion Ratio Lower Upper
225 100
223 59.7 31.9 95.5
227 62.5 32.0 96.2





#95
Naphthalene
Concen: 15.548 ug/l
RT: 15.635 min Scan# 2326
Delta R.T. -0.006 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05




Instrument :
MSVOA_N
ClientSampleId :
VN0905WBS01



Tgt Ion:128 Resp: 12314
Ion Ratio Lower Upper
128 100
127 12.7 10.2 15.2
129 10.9 8.7 13.1

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 09/06/2024
Supervised By :Semsettin Yesilyurt 09/06/2024



#96
1,2,3-Trichlorobenzene
Concen: 14.618 ug/l
RT: 15.841 min Scan# 2361
Delta R.T. 0.000 min
Lab File: VN083668.D
Acq: 05 Sep 2024 12:05

Tgt Ion:180 Resp: 35998
Ion Ratio Lower Upper
180 100
182 97.8 47.1 141.4
145 35.7 17.8 53.5

