

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090618\
 Data File : VN051031.D
 Acq On : 6 Sep 2018 13:24
 Operator : MD\SY
 Sample : J4799-01 5X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EFF-WASTE-WATER

Quant Time: Sep 07 11:18:16 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N082818W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 29 01:38:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	129131	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	675345	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	582868	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.03	65	259934	31.54	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	105.13%
60) 4-Bromofluorobenzene	12.40	95	221458	25.55	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	85.17%
63) Toluene-d8	10.09	98	788625	29.01	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	96.70%

Target Compounds

					Ovalue
15) Acetone	3.82	58	1815489	2600.76	ug/l 96
20) Diisopropyl ether	5.96	45	44492	1.95	ug/l 94
25) Chloroform	7.37	83	183888	13.19	ug/l 99
41) Bromodichloromethane	9.41	83	11469	1.11	ug/l 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090618\
Data File : VN051031.D
Acq On : 6 Sep 2018 13:24
Operator : MD\SY
Sample : J4799-01 5X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EFF-WASTE-WATER

Quant Time: Sep 07 11:18:16 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N082818W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Aug 29 01:38:44 2018
Response via : Initial Calibration

