

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090618\
 Data File : VN051033.D
 Acq On : 6 Sep 2018 15:39
 Operator : MD\SY
 Sample : J4819-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 001M-WILLETS-PT-BLVD(SEP)

Manual Integrations
 APPROVED

MMDadoda
 9/7/2018 11:23:23 AM

Quant Time: Sep 07 07:25:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N082818W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 29 01:38:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.19	128	135484	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.58	114	709479	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	622644	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.03	65	254231	29.41	ug/l	0.00
Spiked Amount 30.000	Range 50 - 169		Recovery =	98.03%		
60) 4-Bromofluorobenzene	12.40	95	250817	27.09	ug/l	0.00
Spiked Amount 30.000	Range 56 - 143		Recovery =	90.30%		
63) Toluene-d8	10.09	98	852458	29.36	ug/l	0.00
Spiked Amount 30.000	Range 66 - 137		Recovery =	97.87%		

Target Compounds

						Ovalue
15) Acetone	3.82	58	1460594	1994.25	ug/l	95
30) 2-Butanone	6.84	43	60961	18.66	ug/l #	95
58) 4-Methyl-2-Pentanone	9.99	43	16052	2.25	ug/l	99
62) Toluene	10.15	91	18355638m	528.42	ug/l	
82) p-Isopropyltoluene	13.29	119	334797	11.88	ug/l	99
91) Naphthalene	15.14	128	6785	8.90	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090618\
Data File : VN051033.D
Acq On : 6 Sep 2018 15:39
Operator : MD\SY
Sample : J4819-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
001M-WILLETS-PT-BLVD(SEP)

Manual Integrations
APPROVED

MMDadoda
9/7/2018 11:23:23 AM

Quant Time: Sep 07 07:25:15 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N082818W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Aug 29 01:38:44 2018
Response via : Initial Calibration

