

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090618\
 Data File : VN051034.D
 Acq On : 6 Sep 2018 16:04
 Operator : MD\SY
 Sample : J4819-01DL 10X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 001M-WILLETS-PT-BLVD(SEP)DL

Quant Time: Sep 07 11:13:46 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N082818W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 29 01:38:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	132012	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	669513	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	586408	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.03	65	254076	30.16	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	100.53%
60) 4-Bromofluorobenzene	12.40	95	208508	23.91	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	79.70%
63) Toluene-d8	10.09	98	792943	29.00	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	96.67%

Target Compounds					Ovalue
15) Acetone	3.82	58	135893	190.42	ug/l 99
62) Toluene	10.16	91	1643750	50.24	ug/l 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090618\
Data File : VN051034.D
Acq On : 6 Sep 2018 16:04
Operator : MD\SY
Sample : J4819-01DL 10X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
001M-WILLETS-PT-BLVD(SEP)DL

Quant Time: Sep 07 11:13:46 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N082818W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Aug 29 01:38:44 2018
Response via : Initial Calibration

