


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN090618\
Data File : VN051027.D
Acq On : 6 Sep 2018 10:55
Operator : MD\SY
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\624N082818W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Wed Aug 29 01:38:44 2018



AutoFind: Scans 3305, 3306, 3307; Background Corrected with Scan 3292

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.3	8473	PASS
75	95	30	60	47.7	26354	PASS
95	95	100	100	100.0	55221	PASS
96	95	5	9	6.0	3286	PASS
173	174	0.00	2	0.4	225	PASS
174	95	50	100	92.7	51216	PASS
175	174	5	9	7.2	3701	PASS
176	174	95	101	97.5	49960	PASS
177	176	5	9	6.9	3452	PASS