

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090619\
 Data File : VN057717.D
 Acq On : 6 Sep 2019 17:09
 Operator : JC/SP
 Sample : PB122820TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 Client Sampled :
 PB122820TB

Manual Integrations
 APPROVED

MMDadoda
 9/9/2019 1:11:03 PM

Quant Time: Sep 06 23:43:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 03:46:54 2019
 Response via : Initial Calibration

| Internal Standards | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|-----------------------------|--------|------|----------|-------|---------|----------|
| 1) Pentafluorobenzene | 7.65 | 168 | 192798 | 50.00 | ug/l | 0.00 |
| 34) 1,4-Difluorobenzene | 8.58 | 114 | 321284 | 50.00 | ug/l | 0.00 |
| 63) Chlorobenzene-d5 | 11.41 | 117 | 294319 | 50.00 | ug/l | 0.00 |
| 72) 1,4-Dichlorobenzene-d4 | 13.35 | 152 | 150611 | 50.00 | ug/l | 0.00 |
| System Monitoring Compounds | | | | | | |
| 33) 1,2-Dichloroethane-d4 | 8.01 | 65 | 246113 | 52.47 | ug/l | 0.00 |
| Spiked Amount | 50.000 | | Recovery | = | 104.94% | |
| 35) Dibromofluoromethane | 7.58 | 113 | 143702 | 49.11 | ug/l | 0.00 |
| Spiked Amount | 50.000 | | Recovery | = | 98.22% | |
| 50) Toluene-d8 | 10.08 | 98 | 568093 | 52.48 | ug/l | 0.00 |
| Spiked Amount | 50.000 | | Recovery | = | 104.96% | |
| 62) 4-Bromofluorobenzene | 12.40 | 95 | 211691 | 47.08 | ug/l | 0.00 |
| Spiked Amount | 50.000 | | Recovery | = | 94.16% | |
| Target Compounds | | | | | | |
| 16) Acetone | 3.81 | 43 | 8577 | 5.942 | ug/l | 92 |
| 20) Methylene Chloride | 4.53 | 84 | 7467m | 2.393 | ug/l | |
| 43) Isopropyl Acetate | 8.15 | 43 | 70974 | 8.714 | ug/l | 98 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090619\
 Data File : VN057717.D
 Acq On : 6 Sep 2019 17:09
 Operator : JC/SP
 Sample : PB122820TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
Client Sampled :
 PB122820TB

Manual Integrations
APPROVED
 MMDadoda
 9/9/2019 1:11:03 PM

Quant Time: Sep 06 23:43:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 03:46:54 2019
 Response via : Initial Calibration

