

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090619\
 Data File : VN057723.D
 Acq On : 6 Sep 2019 19:21
 Operator : JC/SP
 Sample : K4580-21
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T12-A3-(3.5)

Manual Integrations
 APPROVED

MMDadoda
 9/9/2019 1:11:16 PM

Quant Time: Sep 06 23:53:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 03:46:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	192742	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	330272	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	298083	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	152314	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	248226	52.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.88%	
35) Dibromofluoromethane	7.58	113	147616	49.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.14%	
50) Toluene-d8	10.09	98	572248	51.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.84%	
62) 4-Bromofluorobenzene	12.40	95	214864	46.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.98%	
Target Compounds						
16) Acetone	3.81	43	9808m	6.797	ug/l	
43) Isopropyl Acetate	8.15	43	67690	8.085	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN090619\
Data File : VN057723.D
Acq On : 6 Sep 2019 19:21
Operator : JC/SP
Sample : K4580-21
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T12-A3-(3.5)

Manual Integrations
APPROVED

MMDadoda
9/9/2019 1:11:16 PM

Quant Time: Sep 06 23:53:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 03:46:54 2019
Response via : Initial Calibration

