


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN090619\
Data File : VN057709.D
Acq On : 6 Sep 2019 13:53
Operator : JC/SP
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N090619W.M
Title : SW846 8260
Last Update : Fri Sep 06 03:46:54 2019



AutoFind: Scans 3305, 3306, 3307; Background Corrected with Scan 3290

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	25.3	19954	PASS
75	95	30	60	58.7	46232	PASS
95	95	100	100	100.0	78789	PASS
96	95	5	9	6.9	5427	PASS
173	174	0.00	2	1.3	727	PASS
174	95	50	100	70.6	55629	PASS
175	174	5	9	8.0	4475	PASS
176	174	95	101	98.2	54610	PASS
177	176	5	9	6.1	3311	PASS