



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090622\
 Data File : VN074342.D
 Acq On : 06 Sep 2022 15:54
 Operator : JC\MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090622W.M
 Title : SW846 8260
 Last Update : Wed Sep 07 04:58:13 2022



AutoFind: Scans 1832, 1833, 1834; Background Corrected with Scan 1823

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.0	23613	PASS
75	95	30	60	55.4	62152	PASS
95	95	100	100	100.0	112232	PASS
96	95	5	9	6.8	7586	PASS
173	174	0.00	2	1.0	858	PASS
174	95	50	100	75.0	84160	PASS
175	174	5	9	7.9	6622	PASS
176	174	95	101	98.6	82992	PASS
177	176	5	9	6.7	5522	PASS