

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN090718\
 Data File : VN051051.D
 Acq On : 7 Sep 2018 14:11
 Operator : MD\SY
 Sample : J4693-10
 Misc : 5.91g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EP1-MW-C(15-20)

Quant Time: Sep 10 11:10:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N082118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 22 05:38:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	582816	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	875116	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	882229	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	549524	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.02	65	343027	48.89	ug/l	0.00
Spiked Amount			Recovery	=	97.78%	
35) Dibromofluoromethane	7.59	113	326095	46.84	ug/l	0.00
Spiked Amount			Recovery	=	93.68%	
50) Toluene-d8	10.09	98	1431088	54.98	ug/l	0.00
Spiked Amount			Recovery	=	109.96%	
62) 4-Bromofluorobenzene	12.40	95	897383	103.75	ug/l	0.00
Spiked Amount			Recovery	=	207.50%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
39) Methylcyclohexane	9.08	83	252801	22.63	ug/l	99
73) Isopropylbenzene	12.25	105	2787686	61.26	ug/l	100
78) n-propylbenzene	12.59	91	8366497	164.59	ug/l	99
83) tert-Butylbenzene	12.99	119	458006	14.40	ug/l	90
84) 1,2,4-Trimethylbenzene	13.04	105	310635	8.41	ug/l	99
85) sec-Butylbenzene	13.17	105	3440695	82.58	ug/l #	67
89) n-Butylbenzene	13.62	91	2333141	82.17	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090718\
Data File : VN051051.D
Acq On : 7 Sep 2018 14:11
Operator : MD\SY
Sample : J4693-10
Misc : 5.91g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EP1-MW-C(15-20)

Quant Time: Sep 10 11:10:45 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N082118W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 22 05:38:12 2018
Response via : Initial Calibration

