

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN090718\
 Data File : VN051058.D
 Acq On : 7 Sep 2018 17:05
 Operator : MD\SY
 Sample : J4693-10DL 4X
 Misc : 5.91g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EP1-MW-C(15-20)DL

Quant Time: Sep 10 11:44:18 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N082118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 22 05:38:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	661797	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1003252	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	960079	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	460040	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	396505	49.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.54%	
35) Dibromofluoromethane	7.59	113	373896	46.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.70%	
50) Toluene-d8	10.09	98	1465904	49.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.24%	
62) 4-Bromofluorobenzene	12.40	95	591486	59.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	119.30%	
Target Compounds						
39) Methylcyclohexane	9.08	83	49889	3.90	ug/l	98
73) Isopropylbenzene	12.25	105	691445	18.15	ug/l	100
78) n-propylbenzene	12.59	91	2026386	47.62	ug/l	99
83) tert-Butylbenzene	12.99	119	113633	4.27	ug/l	90
85) sec-Butylbenzene	13.17	105	822610	23.58	ug/l #	68
89) n-Butylbenzene	13.62	91	490544	20.64	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN090718\
Data File : VN051058.D
Acq On : 7 Sep 2018 17:05
Operator : MD\SY
Sample : J4693-10DL 4X
Misc : 5.91µ/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EP1-MW-C(15-20)DL

Quant Time: Sep 10 11:44:18 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N082118W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 22 05:38:12 2018
Response via : Initial Calibration

