

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090719\
 Data File : VN057759.D
 Acq On : 7 Sep 2019 17:19
 Operator : JC/SP
 Sample : K4581-31
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T12-D1-(0)

Quant Time: Sep 08 07:01:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 03:46:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	203830	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	328729	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	307578	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	160627	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	235678	47.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.06%	
35) Dibromofluoromethane	7.58	113	139469	46.58	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.16%	
50) Toluene-d8	10.08	98	543509	49.07	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.14%	
62) 4-Bromofluorobenzene	12.40	95	199353	43.33	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.66%	
Target Compounds						
16) Acetone	3.80	43	9021	5.912	ug/l	100
43) Isopropyl Acetate	8.15	43	53681	6.442	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN090719\
Data File : VN057759.D
Acq On : 7 Sep 2019 17:19
Operator : JC/SP
Sample : K4581-31
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T12-D1-(0)

Quant Time: Sep 08 07:01:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 03:46:54 2019
Response via : Initial Calibration

