

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090719\
 Data File : VN057762.D
 Acq On : 7 Sep 2019 18:25
 Operator : JC/SP
 Sample : K4581-34
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T12-D4-(0)

Quant Time: Sep 08 07:10:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 03:46:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	204349	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	317652	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	297049	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	151960	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	242161	48.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.42%	
35) Dibromofluoromethane	7.58	113	142004	49.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.16%	
50) Toluene-d8	10.09	98	555857	51.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.86%	
62) 4-Bromofluorobenzene	12.40	95	199123	44.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.58%	
Target Compounds						
16) Acetone	3.80	43	9002	5.884	ug/l	92
43) Isopropyl Acetate	8.16	43	38288	4.755	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN090719\
Data File : VN057762.D
Acq On : 7 Sep 2019 18:25
Operator : JC/SP
Sample : K4581-34
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T12-D4-(0)

Quant Time: Sep 08 07:10:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 03:46:54 2019
Response via : Initial Calibration

