

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090719\
 Data File : VN057765.D
 Acq On : 7 Sep 2019 19:31
 Operator : JC/SP
 Sample : K4581-37
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T12-D3-(3.5)

Quant Time: Sep 08 07:14:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 03:46:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	205144	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	340787	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	316368	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	162289	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	258259	51.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.50%	
35) Dibromofluoromethane	7.57	113	156302	50.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.70%	
50) Toluene-d8	10.09	98	576071	50.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.34%	
62) 4-Bromofluorobenzene	12.40	95	214379	44.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.90%	
Target Compounds						
16) Acetone	3.80	43	12662	8.245	ug/l	97
43) Isopropyl Acetate	8.16	43	30951	3.583	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN090719\
Data File : VN057765.D
Acq On : 7 Sep 2019 19:31
Operator : JC/SP
Sample : K4581-37
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T12-D3-(3.5)

Quant Time: Sep 08 07:14:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 03:46:54 2019
Response via : Initial Calibration

