

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090719\
 Data File : VN057770.D
 Acq On : 7 Sep 2019 21:21
 Operator : JC/SP
 Sample : K4601-25
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T12-E4-(0)

Quant Time: Sep 08 07:23:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 03:46:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	191110	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	312615	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	286155	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	147724	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	233607	50.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.50%	
35) Dibromofluoromethane	7.58	113	141810	49.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.60%	
50) Toluene-d8	10.08	98	540050	51.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.54%	
62) 4-Bromofluorobenzene	12.40	95	197147	45.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.12%	
Target Compounds						
16) Acetone	3.80	43	11002	7.690	ug/l	99
43) Isopropyl Acetate	8.16	43	36906	4.657	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN090719\
Data File : VN057770.D
Acq On : 7 Sep 2019 21:21
Operator : JC/SP
Sample : K4601-25
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T12-E4-(0)

Quant Time: Sep 08 07:23:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 03:46:54 2019
Response via : Initial Calibration

