

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090719\
 Data File : VN057777.D
 Acq On : 8 Sep 2019 1:01
 Operator : JC/SP
 Sample : K4601-26
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T12-E1-(2.7)

Quant Time: Sep 08 07:26:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 03:46:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	199687	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	322029	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	295777	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	153858	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	239500	49.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.60%	
35) Dibromofluoromethane	7.58	113	145758	49.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.38%	
50) Toluene-d8	10.09	98	542648	50.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.02%	
62) 4-Bromofluorobenzene	12.40	95	200085	44.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.80%	
Target Compounds						
16) Acetone	3.81	43	11088	7.417	ug/l	93
43) Isopropyl Acetate	8.16	43	37548	4.599	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN090719\
Data File : VN057777.D
Acq On : 8 Sep 2019 1:01
Operator : JC/SP
Sample : K4601-26
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T12-E1-(2.7)

Quant Time: Sep 08 07:26:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 03:46:54 2019
Response via : Initial Calibration

