

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090719\
 Data File : VN057760.D
 Acq On : 7 Sep 2019 17:41
 Operator : JC/SP
 Sample : K4581-32
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T12-D2-(0)

Quant Time: Sep 08 07:02:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 03:46:54 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	202544	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	326106	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	298186	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	158242	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	247969	50.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.66%	
35) Dibromofluoromethane	7.57	113	150482	50.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.32%	
50) Toluene-d8	10.09	98	569044	51.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.58%	
62) 4-Bromofluorobenzene	12.40	95	206540	45.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.52%	
Target Compounds						
16) Acetone	3.81	43	9878	6.515	ug/l	95
43) Isopropyl Acetate	8.15	43	43075	5.210	ug/l #	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN090719\
Data File : VN057760.D
Acq On : 7 Sep 2019 17:41
Operator : JC/SP
Sample : K4581-32
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T12-D2-(0)

Quant Time: Sep 08 07:02:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 03:46:54 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.65 min Scan# 1828

Delta R.T. 0.00 min

Lab File: VN057760.D

Acq: 7 Sep 2019 17:41

Instrument :

MSVOA_N

ClientSampleId :


T12-D2-(0)

Tot Ion:168 Resp: 202544

Ion Ratio Lower Upper




168 100

99 68.0 57.4 86.0



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN



#16

Acetone

Concen: 6.515 ug/l

RT: 3.81 min Scan# 632

Delta R.T. 0.01 min

Lab File: VN057760.D


Acq: 7 Sep 2019 17:41

Tgt Ion: 43 Resp: 9878

Ion Ratio Lower Upper



43 100


58 27.4 19.8 29.8



Abundance Ion 43.00 (42.70 to 43.70): VN

Ion 58.00 (57.70 to 58.70): VN









#33
 1,2-Dichloroethane-d4
 Concen: 50.326 ug/l
 RT: 8.01 min Scan# 1941
 Delta R.T. 0.00 min
 Lab File: VN057760.D
 Acq: 7 Sep 2019 17:41




Instrument :
 MSVOA_N
 ClientSampleId :
 T12-D2-(0)


Tot Ion: 65 Resp: 247969
 Ion Ratio Lower Upper
 65 100
 67 49.0 0.0 98.2



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.57 min Scan# 2115
 Delta R.T. 0.00 min
 Lab File: VN057760.D
 Acq: 7 Sep 2019 17:41

Tgt Ion: 114 Resp: 326106
 Ion Ratio Lower Upper
 114 100
 63 25.1 0.0 52.4
 88 16.7 0.0 35.4









#35
 Dibromofluoromethane
 Concen: 50.662 ug/l
 RT: 7.57 min Scan# 1804
 Delta R.T. 0.00 min
 Lab File: VN057760.D
 Acq: 7 Sep 2019 17:41




Instrument :
 MSVOA_N
 ClientSampleId :
 T12-D2-(0)


Tot Ion:113 Resp: 150482
 Ion Ratio Lower Upper
 113 100
 111 101.8 82.7 124.1
 192 18.1 13.8 20.6



#43
 Isopropyl Acetate
 Concen: 5.210 ug/l
 RT: 8.15 min Scan# 1984
 Delta R.T. 0.00 min
 Lab File: VN057760.D
 Acq: 7 Sep 2019 17:41

Tgt Ion: 43 Resp: 43075
 Ion Ratio Lower Upper
 43 100
 61 15.0 12.8 19.2
 87 4.1 7.8 11.8#





#50

Toluene-d8

Concen: 51.788 ug/l

RT: 10.09 min Scan# 2585

Delta R.T. 0.00 min

Lab File: VN057760.D

Acq: 7 Sep 2019 17:41

Instrument :

MSVOA_N

ClientSampleId :


T12-D2-(0)

Tot Ion: 98 Resp: 569044

Ion Ratio Lower Upper

98 100

100 64.1 50.7 76.1



Abundance Ion 98.00 (97.70 to 98.70): VN0

Ion 100.00 (99.70 to 100.70): VN

10.09

300000



200000

100000

0

Time-->

10.00 10.10 10.20



#62

4-Bromofluorobenzene

Concen: 45.255 ug/l

RT: 12.40 min Scan# 3306

Delta R.T. 0.00 min

Lab File: VN057760.D

Acq: 7 Sep 2019 17:41


Tgt Ion: 95 Resp: 206540

Ion Ratio Lower Upper

95 100

174 75.1 0.0 128.0

176 74.3 0.0 128.0



Abundance Ion 94.90 (94.60 to 95.60): VN0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V

12.40

150000


100000


50000

0

Time-->

12.40 12.50





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.41 min Scan# 2996

Delta R.T. 0.00 min

Lab File: VN057760.D

Acq: 7 Sep 2019 17:41

Instrument :

MSVOA_N

ClientSampleId :

T12-D2-(0)


Tot Ion:117 Resp: 298186

Ion Ratio Lower Upper

117 100

82 59.8 51.4 77.0




119 32.6 25.4 38.2



Abundance Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): V

Ion 118.90 (118.60 to 119.60): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.35 min Scan# 3600

Delta R.T. 0.00 min

Lab File: VN057760.D

Acq: 7 Sep 2019 17:41


Tgt Ion:152 Resp: 158242

Ion Ratio Lower Upper

152 100

115 62.3 33.8 101.3

150 155.5 0.0 345.4



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

