

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090721\
 Data File : VN068425.D
 Acq On : 7 Sep 2021 16:03
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Sep 08 11:30:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 08 11:23:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	517606	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	784280	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	746177	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	301778	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	241381	49.769	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.540%
35) Dibromofluoromethane	8.027	113	241522	50.145	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.280%
50) Toluene-d8	10.446	98	924184	52.953	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.900%
62) 4-Bromofluorobenzene	12.733	95	300505	53.738	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	107.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090721\
Data File : VN068425.D
Acq On : 7 Sep 2021 16:03
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Sep 08 11:30:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 08 11:23:05 2021
Response via : Initial Calibration

