

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090723\
 Data File : VN079186.D
 Acq On : 07 Sep 2023 20:07
 Operator : JC\MD
 Sample : 04303-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 230823029-02

Quant Time: Sep 08 05:46:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N090723W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Sep 08 05:09:17 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.824	128	29125	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.112	114	153730	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.871	117	128108	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.582	65	65125	29.762	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	99.200%
60) 4-Bromofluorobenzene	12.853	95	45768	24.247	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	80.833%
63) Toluene-d8	10.571	98	177896	29.714	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	99.033%
Target Compounds						
					Qvalue	
15) Acetone	4.442	58	26857	146.993	ug/l	93
25) Chloroform	7.977	83	36114	10.403	ug/l	99
41) Bromodichloromethane	9.888	83	4384	1.651	ug/l #	85
53) Dibromochloromethane	11.359	129	812	0.414	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090723\
Data File : VN079186.D
Acq On : 07 Sep 2023 20:07
Operator : JC\MD
Sample : 04303-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
230823029-02

Quant Time: Sep 08 05:46:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N090723W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Sep 08 05:09:17 2023
Response via : Initial Calibration

