



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090723\
 Data File : VN079170.D
 Acq On : 07 Sep 2023 08:58
 Operator : JC\MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N090723W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Fri Sep 08 05:09:17 2023



AutoFind: Scans 1852, 1853, 1854; Background Corrected with Scan 1846

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.5	3858	PASS
75	95	30	60	49.5	10895	PASS
95	95	100	100	100.0	22029	PASS
96	95	5	9	6.4	1417	PASS
173	174	0.00	2	0.4	63	PASS
174	95	50	100	78.0	17179	PASS
175	174	5	9	8.3	1428	PASS
176	174	95	101	95.2	16362	PASS
177	176	5	9	6.4	1046	PASS