

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090822\
 Data File : VN074366.D
 Acq On : 08 Sep 2022 18:40
 Operator : JC\MD
 Sample : N4452-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-2-(22)

Quant Time: Sep 09 05:38:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N090822W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Sep 09 05:24:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.592	128	58455	30.000	ug/l	0.01
28) 1,4-Difluorobenzene	8.904	114	337568	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.686	117	297857	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.369	65	192027	31.996	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	106.667%
60) 4-Bromofluorobenzene	12.674	95	152214	25.554	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	85.167%
63) Toluene-d8	10.381	98	496479	29.676	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	98.933%
Target Compounds						
					Qvalue	
25) Chloroform	7.745	83	14437	1.561	ug/l	89
61) Tetrachloroethene	10.916	164	25494	6.899	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090822\
Data File : VN074366.D
Acq On : 08 Sep 2022 18:40
Operator : JC\MD
Sample : N4452-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-2-(22)

Quant Time: Sep 09 05:38:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N090822W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Sep 09 05:24:20 2022
Response via : Initial Calibration

