

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090823\
 Data File : VN079192.D
 Acq On : 08 Sep 2023 12:51
 Operator : JC\MD
 Sample : VN0908WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0908WBL01

Quant Time: Sep 08 17:20:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N090723W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Sep 08 05:09:17 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.818	128	28601	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.106	114	137367	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.871	117	116728	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.583	65	59146	27.525	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	91.733%
60) 4-Bromofluorobenzene	12.859	95	41886	24.353	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	81.167%
63) Toluene-d8	10.571	98	166945	30.603	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	102.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090823\
Data File : VN079192.D
Acq On : 08 Sep 2023 12:51
Operator : JC\MD
Sample : VN0908WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0908WBL01

Quant Time: Sep 08 17:20:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N090723W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Sep 08 05:09:17 2023
Response via : Initial Calibration

