

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090823\
 Data File : VN079194.D
 Acq On : 08 Sep 2023 13:51
 Operator : JC\MD
 Sample : 04320-03RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 001-WILLETS-PT-BLVD(JULY)

Quant Time: Sep 08 17:20:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N090723W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Sep 08 05:09:17 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.818	128	21511	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.106	114	103584	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.871	117	82844	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.588	65	49987	30.930	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	103.100%
60) 4-Bromofluorobenzene	12.853	95	26536	21.739	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	72.467%
63) Toluene-d8	10.571	98	115754	29.898	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	99.667%
Target Compounds						
					Qvalue	
15) Acetone	4.436	58	13092	97.018	ug/l	92
25) Chloroform	7.971	83	88186	34.394	ug/l	99
41) Bromodichloromethane	9.894	83	5318	2.972	ug/l #	95
62) Toluene	10.635	91	80857	16.763	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090823\
Data File : VN079194.D
Acq On : 08 Sep 2023 13:51
Operator : JC\MD
Sample : 04320-03RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
001-WILLETS-PT-BLVD(JULY)

Quant Time: Sep 08 17:20:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N090723W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Sep 08 05:09:17 2023
Response via : Initial Calibration

