

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090823\
 Data File : VN079200.D
 Acq On : 08 Sep 2023 16:17
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Sep 08 23:17:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N090723W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Sep 08 05:09:17 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.818	128	19325	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.112	114	94066	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.870	117	81788	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.588	65	44421	30.595	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	101.967%
60) 4-Bromofluorobenzene	12.853	95	26565	22.044	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	73.467%
63) Toluene-d8	10.571	98	113837	29.783	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	99.267%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090823\
Data File : VN079200.D
Acq On : 08 Sep 2023 16:17
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Sep 08 23:17:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N090723W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Sep 08 05:09:17 2023
Response via : Initial Calibration

