

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN090919\
 Data File : VN057792.D
 Acq On : 9 Sep 2019 11:02
 Operator : JC/SP
 Sample : VSTDICV020
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICSVN090919

Quant Time: Sep 09 12:18:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\624N090919W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Sep 09 12:14:07 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	115	0.00
2 M	Dichlorodifluoromethane	1.813	1.909	-5.3	123	0.00
3 M	Chloromethane	2.260	2.295	-1.5	118	0.00
4 M	Vinyl Chloride	2.312	2.317	-0.2	114	0.00
5 M	Bromomethane	1.443	1.463	-1.4	112	0.00
6 M	Chloroethane	1.437	1.349	6.1	106	0.00
7 M	Trichlorofluoromethane	2.672	2.613	2.2	100	0.00
8 T	Diethyl Ether	1.078	1.128	-4.6	122	0.00
9	1,1,2-Trichlorotrifluoroeth	1.668	1.716	-2.9	119	0.00
10 M	1,1-Dichloroethene	1.586	1.640	-3.4	122	0.00
11	Methyl Iodide	2.122	2.115	0.3	118	0.00
12	Methyl Acetate	2.177	2.151	1.2	114	0.00
13 M	Acrolein	0.118	0.112	5.1	122	0.00
14 M	Acrylonitrile	0.887	0.885	0.2	116	0.00
15 M	Acetone	0.252	0.258	-2.4	111	0.00
16 M	Carbon Disulfide	3.694	3.775	-2.2	126	0.00
17	Allyl chloride	2.907	2.957	-1.7	123	0.00
18 M	Methylene Chloride	1.999	2.022	-1.2	120	0.00
19 M	trans-1,2-Dichloroethene	1.768	1.753	0.8	117	0.00
20 T	Diisopropyl ether	6.796	6.635	2.4	116	0.00
21 M	1,1-Dichloroethane	3.747	3.687	1.6	114	0.00
22 M	cis-1,2-Dichloroethene	2.166	2.209	-2.0	120	0.00
23 M	tert-Butyl Alcohol	0.290	0.293	-1.0	115	0.00
24 M	Methyl tert-Butyl Ether	6.080	6.001	1.3	116	0.00
25 M	Chloroform	3.815	3.757	1.5	116	0.00
26	Cyclohexane	3.148	2.996	4.8	114	0.00
27 s	1,2-Dichloroethane-d4	2.654	2.596	2.2	108	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	111	0.00
29	1,1-Dichloropropene	0.472	0.461	2.3	115	0.00
30 M	2-Butanone	0.213	0.212	0.5	114	0.00
31	2,2-Dichloropropane	0.498	0.493	1.0	121	0.00
32 M	1,1,1-Trichloroethane	0.531	0.509	4.1	113	0.00
33 M	Carbon Tetrachloride	0.433	0.415	4.2	118	0.00
34 M	Benzene	1.398	1.362	2.6	114	0.00
35	Methacrylonitrile	0.197	0.169	14.2	109	0.00
36 M	1,2-Dichloroethane	0.543	0.525	3.3	113	0.00
37 M	Trichloroethene	0.340	0.337	0.9	117	0.00
38	Methylcyclohexane	0.537	0.508	5.4	113	0.00
39 M	1,2-Dichloropropane	0.393	0.374	4.8	113	0.00
40	Dibromomethane	0.239	0.236	1.3	115	0.00
41 M	Bromodichloromethane	0.482	0.455	5.6	116	0.00
42 M	Vinyl Acetate	0.855	0.910	-6.4	117	0.00
43	Ethyl Acetate	0.434	0.425	2.1	112	0.00
44	Isopropyl Acetate	0.757	0.726	4.1	117	0.00
45 T	1,4-Dioxane	0.006	0.006#	0.0	115	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN090919\
 Data File : VN057792.D
 Acq On : 9 Sep 2019 11:02
 Operator : JC/SP
 Sample : VSTDICV020
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleID :
 ICVVN090919

Quant Time: Sep 09 12:18:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\624N090919W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Sep 09 12:14:07 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46	Methyl methacrylate	0.370	0.351	5.1	116	0.00
47	n-amyl Acetate	0.695	0.637	8.3	115	0.00
48 M	t-1,3-Dichloropropene	0.513	0.476	7.2	118	0.00
49 T	cis-1,3-Dichloropropene	0.575	0.542	5.7	117	0.00
50 M	1,1,2-Trichloroethane	0.346	0.329	4.9	111	0.00
51	Ethyl methacrylate	0.537	0.510	5.0	117	0.00
52	1,3-Dichloropropane	0.606	0.586	3.3	114	0.00
53 M	Dibromochloromethane	0.342	0.317	7.3	121	0.00
54 M	1,2-Dibromoethane	0.345	0.326	5.5	114	0.00
55 M	2-Chloroethyl vinyl ether	0.255	0.249	2.4	110	0.00
56 M	Bromoform	0.212	0.189	10.8	121	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	111	0.00
58 M	4-Methyl-2-Pentanone	0.465	0.502	-8.0	116	0.00
59 M	2-Hexanone	0.349	0.361	-3.4	115	0.00
60 S	4-Bromofluorobenzene	0.513	0.500	2.5	111	0.00
61 M	Tetrachloroethene	0.305	0.302	1.0	115	0.00
62 M	Toluene	1.661	1.672	-0.7	114	0.00
63 S	Toluene-d8	1.440	1.424	1.1	108	0.00
64 M	Chlorobenzene	1.030	1.013	1.7	113	0.00
65	1,1,1,2-Tetrachloroethane	0.358	0.346	3.4	117	0.00
66 M	Ethyl Benzene	1.813	1.851	-2.1	114	0.00
67 M	m/p-Xylenes	0.692	0.672	2.9	114	0.00
68 M	o-Xylene	0.678	0.655	3.4	115	0.00
69 M	Styrene	1.156	1.131	2.2	117	0.00
70	Isopropylbenzene	1.792	1.822	-1.7	114	0.00
71 M	1,1,2,2-Tetrachloroethane	0.560	0.548	2.1	115	0.00
72	1,2,3-Trichloropropane	0.497	0.489	1.6	116	0.00
73	Bromobenzene	0.429	0.418	2.6	119	0.00
74	n-propylbenzene	2.048	2.106	-2.8	114	0.00
75	2-Chlorotoluene	1.275	1.249	2.0	116	0.00
76	1,3,5-Trimethylbenzene	1.526	1.525	0.1	115	0.00
77	t-1,4-Dichloro-2-butene	0.147	0.129	12.2	125	0.00
78	4-Chlorotoluene	1.301	1.281	1.5	119	0.00
79	tert-butylbenzene	1.331	1.318	1.0	115	0.00
80	1,2,4-Trimethylbenzene	1.521	1.525	-0.3	115	0.00
81	sec-Butylbenzene	1.766	1.784	-1.0	116	0.00
82	p-Isopropyltoluene	1.582	1.572	0.6	114	0.00
83 M	1,3-Dichlorobenzene	0.798	0.779	2.4	118	0.00
84 M	1,4-Dichlorobenzene	0.785	0.771	1.8	121	0.00
85	n-Butylbenzene	1.474	1.465	0.6	116	0.00
86 T	Hexachloroethane	0.252	0.222	11.9	118	0.00
87 M	1,2-Dichlorobenzene	0.774	0.753	2.7	118	0.00
88	1,2-Dibromo-3-Chloropropane	0.099	0.095	4.0	122	0.00
89	1,2,4-Trichlorobenzene	0.473	0.507	-7.2	135	0.00
90	Hexachlorobutadiene	0.217	0.225	-3.7	123	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN090919\
Data File : VN057792.D
Acq On : 9 Sep 2019 11:02
Operator : JC/SP
Sample : VSTDICV020
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN090919

Quant Time: Sep 09 12:18:53 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\624N090919W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Sep 09 12:14:07 2019
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
91 M	Naphthalene	1.301	1.534	-17.9	143	0.00
92	1,2,3-Trichlorobenzene	0.451	0.504	-11.8	137	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0