

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090920\
 Data File : VN063278.D
 Acq On : 9 Sep 2020 13:28
 Operator : JC/MD
 Sample : L3877-01MEDL 50X
 Misc : 3.84g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-1MEDL

Manual Integrations
 APPROVED

MMDadoda
 9/10/2020 4:04:22 PM

Quant Time: Sep 10 04:09:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 01 04:21:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	447714	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	850550	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	833990	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	401144	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	348458	47.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.22%	
35) Dibromofluoromethane	7.55	113	268495	49.20	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.40%	
50) Toluene-d8	10.06	98	1084748	51.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.30%	
62) 4-Bromofluorobenzene	12.38	95	452722	61.62	ug/l	0.00
Spiked Amount	50.000		Recovery	=	123.24%	
Target Compounds						
40) Benzene	8.00	78	247951	10.319	ug/l	97
67) Ethyl Benzene	11.49	91	906736	29.590	ug/l	98
68) m/p-Xylenes	11.60	106	156886	13.823	ug/l	96
69) o-Xylene	11.92	106	28319	2.583	ug/l	98
73) Isopropylbenzene	12.23	105	40250	1.308	ug/l	92
78) n-propylbenzene	12.57	91	97416	2.736	ug/l	98
80) 1,3,5-Trimethylbenzene	12.71	105	284028	11.091	ug/l	98
84) 1,2,4-Trimethylbenzene	13.02	105	989367	38.328	ug/l	98
86) p-Isopropyltoluene	13.26	119	26098	1.027	ug/l	88
89) n-Butylbenzene	13.59	91	32912m	1.402	ug/l	
95) Naphthalene	15.10	128	7896826	341.532	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN090920\
Data File : VN063278.D
Acq On : 9 Sep 2020 13:28
Operator : JC/MD
Sample : L3877-01MEDL 50X
Misc : 3.84g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-1MEDL

Manual Integrations
APPROVED

MMDadoda
9/10/2020 4:04:22 PM

Quant Time: Sep 10 04:09:35 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 01 04:21:28 2020
Response via : Initial Calibration

