

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090920\
 Data File : VN063280.D
 Acq On : 9 Sep 2020 14:16
 Operator : JC/MD
 Sample : L3870-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SU-04-090820

Manual Integrations
 APPROVED

MMDadoda
 9/10/2020 4:04:26 PM

Quant Time: Sep 10 04:20:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 01 04:21:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	412410	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	786144	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	775139	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	326678	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	327164	48.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.06%	
35) Dibromofluoromethane	7.55	113	247280	49.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.06%	
50) Toluene-d8	10.06	98	999199	51.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.94%	
62) 4-Bromofluorobenzene	12.38	95	394767	58.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.26%	
Target Compounds						
16) Acetone	3.78	43	23962	12.503	ug/l	93
18) Methyl Acetate	4.30	43	5313m	1.081	ug/l	
20) Methylene Chloride	4.50	84	18186	2.912	ug/l #	85
43) Isopropyl Acetate	8.13	43	146372	11.876	ug/l #	90
95) Naphthalene	15.11	128	35556	1.888	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN090920\
Data File : VN063280.D
Acq On : 9 Sep 2020 14:16
Operator : JC/MD
Sample : L3870-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SU-04-090820

Manual Integrations
APPROVED

MMDadoda
9/10/2020 4:04:26 PM

Quant Time: Sep 10 04:20:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 01 04:21:28 2020
Response via : Initial Calibration

