

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090920\
 Data File : VN063284.D
 Acq On : 9 Sep 2020 15:52
 Operator : JC/MD
 Sample : L3932-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VNJ-218

Manual Integrations
 APPROVED

MMDadoda
 9/10/2020 4:04:28 PM

Quant Time: Sep 10 04:30:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 01 04:21:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	407806	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	761769	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	753343	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	321083	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	314765	47.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.44%	
35) Dibromofluoromethane	7.55	113	237811	48.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.32%	
50) Toluene-d8	10.06	98	976771	51.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.86%	
62) 4-Bromofluorobenzene	12.38	95	385631	58.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.22%	
Target Compounds						
16) Acetone	3.78	43	64040	33.793	ug/l	98
18) Methyl Acetate	4.28	43	5018m	1.033	ug/l	
20) Methylene Chloride	4.50	84	19788m	3.254	ug/l	
43) Isopropyl Acetate	8.13	43	140425	11.758	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN090920\
Data File : VN063284.D
Acq On : 9 Sep 2020 15:52
Operator : JC/MD
Sample : L3932-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VNJ-218

Manual Integrations
APPROVED

MMDadoda
9/10/2020 4:04:28 PM

Quant Time: Sep 10 04:30:04 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 01 04:21:28 2020
Response via : Initial Calibration

