

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090921\
 Data File : VN068438.D
 Acq On : 9 Sep 2021 13:19
 Operator : JC/MD
 Sample : VN0909WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0909WBL01

Quant Time: Sep 10 04:07:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 10 04:04:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	466597	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	719960	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	717403	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	273364	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	229726	53.673	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 107.340%	
35) Dibromofluoromethane	8.027	113	229657	52.888	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 105.780%	
50) Toluene-d8	10.446	98	874424	55.437	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 110.880%	
62) 4-Bromofluorobenzene	12.736	95	288706	57.201	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 114.400%	
Target Compounds						
95) Naphthalene	15.512	128	28881	5.979	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090921\
Data File : VN068438.D
Acq On : 9 Sep 2021 13:19
Operator : JC/MD
Sample : VN0909WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0909WBL01

Quant Time: Sep 10 04:07:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 10 04:04:05 2021
Response via : Initial Calibration

