

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090921\
 Data File : VN068442.D
 Acq On : 9 Sep 2021 15:19
 Operator : JC/MD
 Sample : PB138981ZHE#11
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB138981ZHE#11

Quant Time: Sep 10 04:09:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 10 04:04:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	439358	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	708265	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	777341	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	316201	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	235885	58.529	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 117.060%			
35) Dibromofluoromethane	8.027	113	223475	52.315	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 104.620%			
50) Toluene-d8	10.446	98	904195	58.271	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 116.540%			
62) 4-Bromofluorobenzene	12.736	95	317744	63.994	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 127.980%			
Target Compounds					Qvalue	
16) Acetone	4.296	43	65313	48.972	ug/l	97
20) Methylene Chloride	5.100	84	33551	7.756	ug/l	94
25) 2-Butanone	7.343	43	13137	6.708	ug/l	95
43) Isopropyl Acetate	8.563	43	35182	4.440	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090921\
Data File : VN068442.D
Acq On : 9 Sep 2021 15:19
Operator : JC/MD
Sample : PB138981ZHE#11
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB138981ZHE#11

Quant Time: Sep 10 04:09:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 10 04:04:05 2021
Response via : Initial Calibration

