

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090921\
 Data File : VN068443.D
 Acq On : 9 Sep 2021 15:44
 Operator : JC/MD
 Sample : PB138981ZHE#12
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB138981ZHE#12

Quant Time: Sep 10 04:10:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 10 04:04:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	399177	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	650248	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	737784	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	293507	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	218669	59.719	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 119.440%			
35) Dibromofluoromethane	8.024	113	206433	52.637	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 105.280%			
50) Toluene-d8	10.446	98	843190	59.188	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 118.380%			
62) 4-Bromofluorobenzene	12.734	95	298411	65.462	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 130.920%			
Target Compounds						Qvalue
16) Acetone	4.296	43	64794	53.474	ug/l	94
20) Methylene Chloride	5.103	84	31912	8.187	ug/l	98
25) 2-Butanone	7.348	43	13508	7.592	ug/l	100
43) Isopropyl Acetate	8.563	43	37416	5.143	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090921\
Data File : VN068443.D
Acq On : 9 Sep 2021 15:44
Operator : JC/MD
Sample : PB138981ZHE#12
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB138981ZHE#12

Quant Time: Sep 10 04:10:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 10 04:04:05 2021
Response via : Initial Calibration

