

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090921\
 Data File : VN068444.D
 Acq On : 9 Sep 2021 16:09
 Operator : JC/MD
 Sample : PB138981ZHE#13
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB138981ZHE#13

Quant Time: Sep 10 04:10:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 10 04:04:05 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	397691	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	652203	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	733174	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	286877	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	221060	60.597	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	121.200%
35) Dibromofluoromethane	8.024	113	204263	51.927	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.860%
50) Toluene-d8	10.446	98	845979	59.206	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	118.420%
62) 4-Bromofluorobenzene	12.736	95	296643	64.879	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	129.760%
Target Compounds						
						Qvalue
16) Acetone	4.291	43	39086	32.378	ug/l	91
20) Methylene Chloride	5.100	84	59343	16.511	ug/l	98
43) Isopropyl Acetate	8.566	43	39671	5.436	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090921\
Data File : VN068444.D
Acq On : 9 Sep 2021 16:09
Operator : JC/MD
Sample : PB138981ZHE#13
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB138981ZHE#13

Quant Time: Sep 10 04:10:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 10 04:04:05 2021
Response via : Initial Calibration

