

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090921\
 Data File : VN068446.D
 Acq On : 9 Sep 2021 17:00
 Operator : JC/MD
 Sample : M3670-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SAND-BLAST-MEDIA

Quant Time: Sep 10 04:11:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 10 04:04:05 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	453509	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	723331	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	795958	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	404398	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	229290	55.117	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	110.240%
35) Dibromofluoromethane	8.024	113	222925	51.099	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.200%
50) Toluene-d8	10.446	98	904784	57.095	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	114.180%
62) 4-Bromofluorobenzene	12.734	95	346019	68.237	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	136.480%#
Target Compounds						
						Qvalue
16) Acetone	4.293	43	81757	59.389	ug/l	97
20) Methylene Chloride	5.101	84	33661	7.499	ug/l	94
25) 2-Butanone	7.343	43	13563	6.709	ug/l	97
43) Isopropyl Acetate	8.568	43	41088	5.077	ug/l #	83
95) Naphthalene	15.509	128	1043585	64.126	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090921\
Data File : VN068446.D
Acq On : 9 Sep 2021 17:00
Operator : JC/MD
Sample : M3670-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SAND-BLAST-MEDIA

Quant Time: Sep 10 04:11:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 10 04:04:05 2021
Response via : Initial Calibration

