

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090921\
 Data File : VN068447.D
 Acq On : 9 Sep 2021 17:25
 Operator : JC/MD
 Sample : M3671-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FILTER-MEDIA-FILTERS

Quant Time: Sep 10 04:11:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 10 04:04:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	408133	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	676486	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	734365	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	358926	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	218607	58.392	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 116.780%			
35) Dibromofluoromethane	8.027	113	204573	50.139	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 100.280%			
50) Toluene-d8	10.446	98	850539	57.388	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 114.780%			
62) 4-Bromofluorobenzene	12.733	95	323813	68.280	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 136.560%#			
Target Compounds						
					Qvalue	
16) Acetone	4.291	43	76638	61.860	ug/l	99
20) Methylene Chloride	5.101	84	23291	5.438	ug/l	95
25) 2-Butanone	7.345	43	25036	13.762	ug/l	94
43) Isopropyl Acetate	8.566	43	36406	4.810	ug/l #	81
95) Naphthalene	15.509	128	263586	20.750	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090921\
Data File : VN068447.D
Acq On : 9 Sep 2021 17:25
Operator : JC/MD
Sample : M3671-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FILTER-MEDIA-FILTERS

Quant Time: Sep 10 04:11:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 10 04:04:05 2021
Response via : Initial Calibration

