

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090921\
 Data File : VN068449.D
 Acq On : 9 Sep 2021 18:16
 Operator : JC/MD
 Sample : M3670-01RE
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SAND-BLAST-MEDIARE

Manual Integrations
 APPROVED

MMDadoda
 9/10/2021 12:47:15 PM

Quant Time: Sep 10 04:12:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 10 04:04:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	479265	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	755499	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	844952	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	435962	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	242271	55.108	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 110.220%			
35) Dibromofluoromethane	8.024	113	231727	50.855	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 101.700%			
50) Toluene-d8	10.446	98	943394	56.996	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 114.000%			
62) 4-Bromofluorobenzene	12.733	95	362002	68.349	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 136.700%#			
Target Compounds					Qvalue	
16) Acetone	4.293	43	87442	60.105	ug/l	98
20) Methylene Chloride	5.109	84	34515m	7.234	ug/l	
25) 2-Butanone	7.340	43	14482	6.779	ug/l	95
43) Isopropyl Acetate	8.563	43	46344	5.482	ug/l #	83
95) Naphthalene	15.507	128	1244939	70.587	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090921\
Data File : VN068449.D
Acq On : 9 Sep 2021 18:16
Operator : JC/MD
Sample : M3670-01RE
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SAND-BLAST-MEDIARE

Manual Integrations
APPROVED

MMDadoda
9/10/2021 12:47:15 PM

Quant Time: Sep 10 04:12:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 10 04:04:05 2021
Response via : Initial Calibration

