

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090922\
 Data File : VN074375.D
 Acq On : 09 Sep 2022 13:13
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Sep 10 06:13:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 06:10:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	340989	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	592392	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	561580	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	244958	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	242071	41.032	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	82.060%
35) Dibromofluoromethane	7.945	113	194533	48.643	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.280%
50) Toluene-d8	10.380	98	678451	43.538	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	87.080%
62) 4-Bromofluorobenzene	12.674	95	271806	48.043	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.080%
Target Compounds						Qvalue
75) 1,1,2,2-Tetrachloroethane	12.769	83	1882	0.213	ug/l #	83
77) Bromobenzene	12.810	156	1664	0.329	ug/l	87
78) n-propylbenzene	12.869	91	6173	0.229	ug/l	95
79) 2-Chlorotoluene	12.945	91	3584	0.218	ug/l	88
82) 4-Chlorotoluene	13.045	91	6160	0.388	ug/l	100
86) p-Isopropyltoluene	13.557	119	3894	0.203	ug/l #	85
87) 1,3-Dichlorobenzene	13.557	146	3895	0.417	ug/l	95
89) n-Butylbenzene	13.886	91	4833	0.297	ug/l	95
91) 1,2-Dichlorobenzene	13.933	146	3551	0.361	ug/l	92
93) 1,2,4-Trichlorobenzene	15.198	180	7057	1.442	ug/l	80
95) Naphthalene	15.445	128	35939	1.877	ug/l	97
96) 1,2,3-Trichlorobenzene	15.627	180	6525	1.288	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090922\
Data File : VN074375.D
Acq On : 09 Sep 2022 13:13
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Sep 10 06:13:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 10 06:10:29 2022
Response via : Initial Calibration

