

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090922\
 Data File : VN074384.D
 Acq On : 09 Sep 2022 19:45
 Operator : JC\MD
 Sample : N4471-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-11A-11-16-083022-FD

Quant Time: Sep 10 07:25:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 06:10:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	367754	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	641722	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	624443	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	296377	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	259126	40.726	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	81.460%
35) Dibromofluoromethane	7.951	113	218106	50.345	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.680%
50) Toluene-d8	10.380	98	736657	43.640	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	87.280%
62) 4-Bromofluorobenzene	12.674	95	300664	49.058	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.120%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	2.404	62	557578	70.029	ug/l	95
9) 1,1,2-Trichlorotrifluo...	4.140	101	349234	74.129	ug/l	98
12) 1,1-Dichloroethene	4.122	96	25030	5.460	ug/l	94
21) trans-1,2-Dichloroethene	5.528	96	20068	3.907	ug/l	96
24) 1,1-Dichloroethane	6.304	63	32036	3.079	ug/l #	85
27) cis-1,2-Dichloroethene	7.251	96	7333276	1158.715	ug/l	90
29) Tetrahydrofuran	7.616	42	18196	4.774	ug/l	98
30) Chloroform	7.745	83	2838	0.273	ug/l #	70
40) Benzene	8.398	78	12320	0.564	ug/l	95
42) 1,2-Dichloroethane	8.463	62	3078	0.368	ug/l	97
44) Trichloroethene	9.151	130	8148744	1663.346	ug/l	100
55) 1,1,2-Trichloroethane	10.839	97	2709	0.486	ug/l #	88
64) Tetrachloroethene	10.922	164	45838	10.828	ug/l	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090922\
Data File : VN074384.D
Acq On : 09 Sep 2022 19:45
Operator : JC\MD
Sample : N4471-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-11A-11-16-083022-FD

Quant Time: Sep 10 07:25:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 10 06:10:29 2022
Response via : Initial Calibration

