

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090922\
 Data File : VN074386.D
 Acq On : 09 Sep 2022 20:34
 Operator : JC\MD
 Sample : N4471-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-12A-13-18-083122

Quant Time: Sep 10 07:26:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 06:10:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	347647	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	631282	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	602662	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	276646	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	250100	41.581	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	83.160%
35) Dibromofluoromethane	7.951	113	207264	48.633	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.260%
50) Toluene-d8	10.380	98	701323	42.234	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	84.460%#
62) 4-Bromofluorobenzene	12.674	95	291653	48.375	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.760%
Target Compounds						
27) cis-1,2-Dichloroethene	7.251	96	18650	3.117	ug/l	88
44) Trichloroethene	9.151	130	79002	16.393	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090922\
Data File : VN074386.D
Acq On : 09 Sep 2022 20:34
Operator : JC\MD
Sample : N4471-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-12A-13-18-083122

Quant Time: Sep 10 07:26:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 10 06:10:29 2022
Response via : Initial Calibration

