

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090922\
 Data File : VN074389.D
 Acq On : 09 Sep 2022 21:47
 Operator : JC\MD
 Sample : N4471-13
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-12B-45-50-083122

Quant Time: Sep 10 07:27:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 06:10:29 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 09/12/2022
 Supervised By : Mahesh Dadoda 09/12/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	335710	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	592331	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	586777	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	264713	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	240354	41.381	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	82.760%		
35) Dibromofluoromethane	7.951	113	196268	49.082	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	98.160%		
50) Toluene-d8	10.380	98	674722	43.304	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	86.600%		
62) 4-Bromofluorobenzene	12.674	95	277024	48.970	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	97.940%		
Target Compounds						
21) trans-1,2-Dichloroethene	5.516	96	5132m	1.094	ug/l	Qvalue
27) cis-1,2-Dichloroethene	7.245	96	104163	18.030	ug/l	91
44) Trichloroethene	9.151	130	40196	8.889	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090922\
Data File : VN074389.D
Acq On : 09 Sep 2022 21:47
Operator : JC\MD
Sample : N4471-13
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-12B-45-50-083122

Quant Time: Sep 10 07:27:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 10 06:10:29 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 09/12/2022
Supervised By :Mahesh Dadoda 09/12/2022

