

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090922\
 Data File : VN074391.D
 Acq On : 09 Sep 2022 22:36
 Operator : JC\MD
 Sample : N4471-17
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB01-083122

Quant Time: Sep 10 07:28:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 06:10:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	333195	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	593422	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	579506	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	267882	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	241266	41.852	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	83.700%
35) Dibromofluoromethane	7.951	113	201253	50.236	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.480%
50) Toluene-d8	10.380	98	668100	42.800	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	85.600%#
62) 4-Bromofluorobenzene	12.674	95	275678	48.643	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.280%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	4.140	101	60207	14.105	ug/l	99
16) Acetone	4.222	43	28274	9.085	ug/l	96
27) cis-1,2-Dichloroethene	7.251	96	10954	1.910	ug/l	89
44) Trichloroethene	9.151	130	545888	120.498	ug/l	97
64) Tetrachloroethene	10.910	164	3964	1.009	ug/l #	60

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090922\
Data File : VN074391.D
Acq On : 09 Sep 2022 22:36
Operator : JC\MD
Sample : N4471-17
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB01-083122

Quant Time: Sep 10 07:28:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 10 06:10:29 2022
Response via : Initial Calibration

