

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090922\
 Data File : VN074393.D
 Acq On : 09 Sep 2022 23:25
 Operator : JC\MD
 Sample : N4509-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S1-SO_112D-20220902

Quant Time: Sep 10 07:28:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 06:10:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	344006	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	616537	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	634406	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	288258	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	243282	40.875	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	81.760%
35) Dibromofluoromethane	7.951	113	201422	48.393	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.780%
50) Toluene-d8	10.380	98	717528	44.243	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	88.480%
62) 4-Bromofluorobenzene	12.674	95	294297	49.981	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.960%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	4.157	101	13375	3.035	ug/l	99
16) Acetone	4.228	43	58298	18.143	ug/l	95
20) Methylene Chloride	5.016	84	18109	2.409	ug/l	94
37) Ethyl Acetate	7.334	43	81201	8.163	ug/l	99
43) Isopropyl Acetate	8.498	43	333446	20.498	ug/l #	91
44) Trichloroethene	9.151	130	94116	19.996	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090922\
Data File : VN074393.D
Acq On : 09 Sep 2022 23:25
Operator : JC\MD
Sample : N4509-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S1-SO_112D-20220902

Quant Time: Sep 10 07:28:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 10 06:10:29 2022
Response via : Initial Calibration

