

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090922\
 Data File : VN074397.D
 Acq On : 10 Sep 2022 01:02
 Operator : JC\MD
 Sample : N4509-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S1-SO_PW001-20220902

Quant Time: Sep 10 07:30:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 06:10:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	312118	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	554951	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	588273	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	266980	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	226422	41.929	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	83.860%
35) Dibromofluoromethane	7.951	113	182841	48.804	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.600%
50) Toluene-d8	10.380	98	646311	44.274	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	88.540%
62) 4-Bromofluorobenzene	12.674	95	274501	51.793	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.580%
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	59886	20.541	ug/l	96
18) Methyl Acetate	4.787	43	10067	1.171	ug/l #	88
20) Methylene Chloride	5.010	84	19103	3.053	ug/l	88
37) Ethyl Acetate	7.339	43	76029	8.492	ug/l	99
43) Isopropyl Acetate	8.492	43	341844	23.347	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090922\
Data File : VN074397.D
Acq On : 10 Sep 2022 01:02
Operator : JC\MD
Sample : N4509-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S1-SO_PW001-20220902

Quant Time: Sep 10 07:30:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 10 06:10:29 2022
Response via : Initial Calibration

