

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090922\
 Data File : VN074399.D
 Acq On : 10 Sep 2022 01:51
 Operator : JC\MD
 Sample : N4502-02 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DEACON-WATER-COMP

Quant Time: Sep 10 07:31:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 06:10:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	319374	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	565056	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	548912	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	253495	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	229232	41.485	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	82.980%
35) Dibromofluoromethane	7.951	113	189523	49.683	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.360%
50) Toluene-d8	10.380	98	642203	43.206	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	86.420%
62) 4-Bromofluorobenzene	12.674	95	258989	47.992	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090922\
Data File : VN074399.D
Acq On : 10 Sep 2022 01:51
Operator : JC\MD
Sample : N4502-02 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DEACON-WATER-COMP

Quant Time: Sep 10 07:31:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 10 06:10:29 2022
Response via : Initial Calibration

