


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090922\
Data File : VN074368.D
Acq On : 09 Sep 2022 10:01
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
Title : SW846 8260
Last Update : Sat Sep 10 06:10:29 2022



AutoFind: Scans 1833, 1834, 1835; Background Corrected with Scan 1825

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.4	17637	PASS
75	95	30	60	56.6	42693	PASS
95	95	100	100	100.0	75461	PASS
96	95	5	9	5.7	4271	PASS
173	174	0.00	2	0.4	217	PASS
174	95	50	100	69.8	52693	PASS
175	174	5	9	8.0	4228	PASS
176	174	95	101	99.5	52419	PASS
177	176	5	9	6.2	3243	PASS