

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090924\
 Data File : VN083708.D
 Acq On : 09 Sep 2024 10:03
 Operator : JC\MD
 Sample : VN0909WBSD01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Sep 10 01:16:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N090624W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Sep 06 23:42:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.812	128	31032	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.094	114	176340	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.865	117	157868	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.577	65	84062	29.635	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	98.800%
60) 4-Bromofluorobenzene	12.847	95	81142	29.823	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	99.400%
63) Toluene-d8	10.565	98	219979	30.093	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.300%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.124	85	35706	17.182	ug/l	96
3) Chloromethane	2.359	50	37868	17.538	ug/l	94
4) Vinyl Chloride	2.512	62	38226	17.221	ug/l	91
5) Bromomethane	2.901	94	21688	17.488	ug/l	93
6) Chloroethane	3.071	64	27267	18.344	ug/l	92
7) Trichlorofluoromethane	3.465	101	64048	17.211	ug/l	98
8) Diethyl Ether	3.954	74	22619	17.888	ug/l	84
9) 1,1,2-Trichlorotrifluo...	4.353	101	21528	10.770	ug/l	47
10) 1,1-Dichloroethene	4.324	96	32968	17.093	ug/l #	74
11) Methyl Iodide	4.571	142	44708	17.066	ug/l	92
12) Methyl Acetate	5.018	43	43491	18.698	ug/l #	77
13) Acrolein	4.177	56	36591	107.792	ug/l	98
14) Acrylonitrile	5.712	53	84766	90.052	ug/l	95
15) Acetone	4.430	58	28466	91.165	ug/l	75
16) Carbon Disulfide	4.701	76	90091	16.528	ug/l	96
17) Allyl chloride	5.012	41	57665	17.050	ug/l	87
18) Methylene Chloride	5.265	84	36879	17.477	ug/l	92
19) trans-1,2-Dichloroethene	5.777	96	34686	17.352	ug/l	97
20) Diisopropyl ether	6.671	45	118820	17.396	ug/l	96
21) 1,1-Dichloroethane	6.565	63	68303	17.515	ug/l	97
22) cis-1,2-Dichloroethene	7.483	96	41012	16.857	ug/l	92
23) tert-Butyl Alcohol	5.536	59	35151	101.764	ug/l #	100
24) Methyl tert-Butyl Ether	5.795	73	121305	17.699	ug/l #	87
25) Chloroform	7.965	83	72159	17.460	ug/l	99
26) Cyclohexane	8.253	56	57616	16.775	ug/l #	89
29) 1,1-Dichloropropene	8.371	75	48120	17.204	ug/l	98
30) 2-Butanone	7.483	43	125228	94.980	ug/l	96
31) 2,2-Dichloropropane	7.489	77	64997	18.227	ug/l	97
32) 1,1,1-Trichloroethane	8.165	97	66397	17.775	ug/l	98
33) Carbon Tetrachloride	8.359	117	58022	17.937	ug/l	99
34) Benzene	8.600	78	151159	18.012	ug/l #	95
35) Methacrylonitrile	7.777	41	29187	19.692	ug/l	91
36) 1,2-Dichloroethane	8.671	62	57105	18.100	ug/l	95
37) Trichloroethene	9.353	130	35056	17.789	ug/l	94
38) Methylcyclohexane	9.594	83	59286	17.705	ug/l	97
39) 1,2-Dichloropropane	9.618	63	36311	17.978	ug/l	97
40) Dibromomethane	9.706	93	26610	18.018	ug/l	94
41) Bromodichloromethane	9.888	83	55660	17.723	ug/l	98
42) Vinyl Acetate	6.600	43	450676	83.137	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090924\
 Data File : VN083708.D
 Acq On : 09 Sep 2024 10:03
 Operator : JC\MD
 Sample : VN0909WBSD01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Sep 10 01:16:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N090624W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Sep 06 23:42:47 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) Ethyl Acetate	7.559	43	48775	18.379	ug/l	97
44) Isopropyl Acetate	8.683	43	85971	18.730	ug/l	90
45) 1,4-Dioxane	9.694	88	15043	420.192	ug/l	98
46) Methyl methacrylate	9.683	41	42453	19.109	ug/l	80
47) n-amyl Acetate	12.494	43	77395	19.035	ug/l #	89
48) t-1,3-Dichloropropene	10.835	75	57570	17.428	ug/l	96
49) cis-1,3-Dichloropropene	10.312	75	61875	18.011	ug/l #	87
50) 1,1,2-Trichloroethane	11.012	97	35316	18.793	ug/l	96
51) Ethyl methacrylate	10.871	69	58165	18.117	ug/l	70
52) 1,3-Dichloropropane	11.159	76	61067	17.994	ug/l	99
53) Dibromochloromethane	11.359	129	40918	18.153	ug/l	96
54) 1,2-Dibromoethane	11.471	107	35213	18.113	ug/l	99
55) 2-Chloroethyl vinyl ether	10.159	63	109480	85.564	ug/l	99
56) Bromoform	12.576	173	25863	18.061	ug/l	99
58) 4-Methyl-2-Pentanone	10.441	43	248553	99.295	ug/l #	88
59) 2-Hexanone	11.194	43	189895	100.397	ug/l	88
61) Tetrachloroethene	11.100	164	28874	17.282	ug/l	95
62) Toluene	10.630	91	160730	17.713	ug/l	100
64) Chlorobenzene	11.888	112	98169	17.474	ug/l	98
65) 1,1,1,2-Tetrachloroethane	11.959	131	35072	17.921	ug/l	95
66) Ethyl Benzene	11.965	91	177087	17.161	ug/l	96
67) m/p-Xylenes	12.071	106	131117	34.537	ug/l	96
68) o-Xylene	12.394	106	64783	17.632	ug/l	95
69) Styrene	12.412	104	106951	17.227	ug/l	98
70) Isopropylbenzene	12.694	105	171180	17.654	ug/l	99
71) 1,1,2,2-Tetrachloroethane	12.935	83	50581	19.099	ug/l	99
72) 1,2,3-Trichloropropane	12.994	75	67704	29.158	ug/l	64
73) Bromobenzene	12.976	156	39441	17.709	ug/l	94
74) n-propylbenzene	13.035	91	194312	17.119	ug/l	100
75) 2-Chlorotoluene	13.124	91	125200	17.673	ug/l	99
76) 1,3,5-Trimethylbenzene	13.171	105	143082	17.665	ug/l	98
77) t-1,4-Dichloro-2-butene	12.735	75	16943	17.263	ug/l	87
78) 4-Chlorotoluene	13.223	91	120926	16.832	ug/l	99
79) tert-butylbenzene	13.435	119	126257	17.826	ug/l	98
80) 1,2,4-Trimethylbenzene	13.482	105	143312	17.509	ug/l	99
81) sec-Butylbenzene	13.618	105	164949	17.048	ug/l	100
82) p-Isopropyltoluene	13.729	119	138870	17.059	ug/l	97
83) 1,3-Dichlorobenzene	13.729	146	72292	16.865	ug/l	98
84) 1,4-Dichlorobenzene	13.812	146	71460	16.675	ug/l	99
85) n-Butylbenzene	14.053	91	117524	16.360	ug/l	98
86) Hexachloroethane	14.335	117	26943	17.465	ug/l	88
87) 1,2-Dichlorobenzene	14.106	146	71564	17.203	ug/l	99
88) 1,2-Dibromo-3-Chloropr...	14.718	75	10930	18.780	ug/l	93
89) 1,2,4-Trichlorobenzene	15.841	180	36209	15.888	ug/l	99
90) Hexachlorobutadiene	15.500	225	16072	17.038	ug/l	99
91) Naphthalene	15.641	128	122921	16.498	ug/l	99
92) 1,2,3-Trichlorobenzene	15.841	180	36209	15.888	ug/l	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090924\
 Data File : VN083708.D
 Acq On : 09 Sep 2024 10:03
 Operator : JC\MD
 Sample : VN0909WBSD01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Sep 10 01:16:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N090624W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Sep 06 23:42:47 2024
 Response via : Initial Calibration









#1
Bromochloromethane
Concen: 30.000 ug/l
RT: 7.812 min Scan# 996
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:128 Resp: 31032
Ion Ratio Lower Upper
128 100
49 173.1 0.0 403.7
130 125.8 0.0 323.3



#2
Dichlorodifluoromethane
Concen: 17.182 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03

Tgt Ion: 85 Resp: 35706
Ion Ratio Lower Upper
85 100
87 31.0 16.8 50.3





#3

Chloromethane

Concen: 17.538 ug/l

RT: 2.359 min Scan# 69

Delta R.T. 0.000 min

Lab File: VN083708.D

Acq: 09 Sep 2024 10:03

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 50 Resp: 37868

Ion Ratio Lower Upper

50 100

52 34.8 25.4 38.0



#4

Vinyl Chloride

Concen: 17.221 ug/l

RT: 2.512 min Scan# 95

Delta R.T. 0.000 min

Lab File: VN083708.D




Acq: 09 Sep 2024 10:03


Tgt Ion: 62 Resp: 38226

Ion Ratio Lower Upper

62 100

64 35.1 24.1 36.1









#5
Bromomethane
Concen: 17.488 ug/l
RT: 2.901 min Scan# 1
Delta R.T. -0.006 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 94 Resp: 21688
Ion Ratio Lower Upper
94 100
96 84.5 72.8 109.2



#6
Chloroethane
Concen: 18.344 ug/l
RT: 3.071 min Scan# 190
Delta R.T. -0.012 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03


Tgt Ion: 64 Resp: 27267
Ion Ratio Lower Upper
64 100
66 28.8 26.7 40.1








#7
 Trichlorofluoromethane
 Concen: 17.211 ug/l
 RT: 3.465 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: VN083708.D
 Acq: 09 Sep 2024 10:03

Instrument :
 MSVOA_N
 ClientSampleId :







Tgt Ion:101 Resp: 64048
 Ion Ratio Lower Upper
 101 100
 103 64.3 53.0 79.4



#8
 Diethyl Ether
 Concen: 17.888 ug/l
 RT: 3.954 min Scan# 340
 Delta R.T. 0.000 min
 Lab File: VN083708.D
 Acq: 09 Sep 2024 10:03

Tgt Ion: 74 Resp: 22619
 Ion Ratio Lower Upper
 74 100
 45 96.6 16.4 147.4









#9
1,1,2-Trichlorotrifluoroethane
Concen: 10.770 ug/l
RT: 4.353 min Scan# 409
Delta R.T. -0.006 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:101 Resp: 21528
Ion Ratio Lower Upper
101 100
85 81.3 0.0 90.6
151 124.0 0.0 157.6



#10
1,1-Dichloroethene
Concen: 17.093 ug/l
RT: 4.324 min Scan# 403
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03


Tgt Ion: 96 Resp: 32968
Ion Ratio Lower Upper
96 100
61 183.9 117.6 176.4#
98 73.4 48.4 72.6#








#11
Methyl Iodide
Concen: 17.066 ug/l
RT: 4.571 min Scan# 447
Delta R.T. -0.012 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion:142 Resp: 44708
Ion Ratio Lower Upper
142 100
127 47.2 26.3 78.9
141 13.4 8.3 24.9



#12
Methyl Acetate
Concen: 18.698 ug/l
RT: 5.018 min Scan# 521
Delta R.T. -0.006 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03

Tgt Ion: 43 Resp: 43491
Ion Ratio Lower Upper
43 100
74 21.4 27.7 41.5#





#13

Acrolein

Concen: 107.792 ug/l

RT: 4.177 min Scan# 31

Delta R.T. 0.000 min

Lab File: VN083708.D

Acq: 09 Sep 2024 10:03

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 56 Resp: 36591

Ion Ratio Lower Upper

56 100

55 70.9 55.5 83.3



#14

Acrylonitrile

Concen: 90.052 ug/l

RT: 5.712 min Scan# 639

Delta R.T. 0.000 min

Lab File: VN083708.D

Acq: 09 Sep 2024 10:03




Tgt Ion: 53 Resp: 84766


Ion Ratio Lower Upper

53 100

52 86.7 41.3 123.9

51 37.0 16.8 50.3





#15

Acetone

Concen: 91.165 ug/l

RT: 4.430 min Scan# 41

Delta R.T. 0.000 min

Lab File: VN083708.D

Acq: 09 Sep 2024 10:03

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 58 Resp: 28466

Ion Ratio Lower Upper

58 100

43 325.4 223.1 334.7



#16

Carbon Disulfide

Concen: 16.528 ug/l

RT: 4.701 min Scan# 467

Delta R.T. 0.000 min

Lab File: VN083708.D




Acq: 09 Sep 2024 10:03


Tgt Ion: 76 Resp: 90091

Ion Ratio Lower Upper

76 100

78 9.4 6.3 9.5





#17

Allyl chloride

Concen: 17.050 ug/l

RT: 5.012 min Scan# 51

Delta R.T. 0.000 min


Lab File: VN083708.D

Acq: 09 Sep 2024 10:03

Instrument :

MSVOA_N

ClientSampleId :






Tgt Ion: 41 Resp: 57665

Ion Ratio Lower Upper

41 100

39 77.8 37.6 112.9

76 36.7 27.8 83.4



#18

Methylene Chloride

Concen: 17.477 ug/l

RT: 5.265 min Scan# 563

Delta R.T. 0.000 min

Lab File: VN083708.D

Acq: 09 Sep 2024 10:03

Tgt Ion: 84 Resp: 36879




Ion Ratio Lower Upper


84 100

49 115.5 86.0 129.0

51 38.5 30.0 45.0


86 55.8 53.0 79.4








#19
trans-1,2-Dichloroethene
Concen: 17.352 ug/l
RT: 5.777 min Scan# 610
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03


Instrument :
MSVOA_N
ClientSampleId :





Tgt Ion: 96 Resp: 34686
Ion Ratio Lower Upper
96 100
61 137.5 111.8 167.6
98 58.1 49.2 73.8




#20
Diisopropyl ether
Concen: 17.396 ug/l
RT: 6.671 min Scan# 802
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03



Tgt Ion: 45 Resp: 118820
Ion Ratio Lower Upper
45 100
43 53.2 40.4 60.6
87 27.5 23.8 35.8
59 12.1 9.8 14.6









#21
1,1-Dichloroethane
Concen: 17.515 ug/l
RT: 6.565 min Scan# 71
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 63 Resp: 68303
Ion Ratio Lower Upper
63 100
98 8.8 4.1 12.3
100 4.5 3.0 9.0



#22
cis-1,2-Dichloroethene
Concen: 16.857 ug/l
RT: 7.483 min Scan# 940
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03


Tgt Ion: 96 Resp: 41012
Ion Ratio Lower Upper
96 100
61 155.7 113.5 170.3
98 66.6 52.0 78.0








#23
 tert-Butyl Alcohol
 Concen: 101.764 ug/l
 RT: 5.536 min Scan# 609
 Delta R.T. 0.000 min
 Lab File: VN083708.D
 Acq: 09 Sep 2024 10:03

Instrument :
 MSVOA_N
 ClientSampleId :







Tgt Ion: 59 Resp: 35151
 Ion Ratio Lower Upper
 59 100
 52 0.0 0.0 0.0



#24
 Methyl tert-Butyl Ether
 Concen: 17.699 ug/l
 RT: 5.795 min Scan# 653
 Delta R.T. 0.006 min
 Lab File: VN083708.D
 Acq: 09 Sep 2024 10:03

Tgt Ion: 73 Resp: 121305
 Ion Ratio Lower Upper
 73 100
 43 22.4 13.4 20.2#





#25

Chloroform

Concen: 17.460 ug/l

RT: 7.965 min Scan# 1022

Delta R.T. 0.000 min

Lab File: VN083708.D

Acq: 09 Sep 2024 10:03

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 83 Resp: 72159

Ion Ratio Lower Upper

83 100

85 64.3 50.9 76.3



#26

Cyclohexane

Concen: 16.775 ug/l

RT: 8.253 min Scan# 1071

Delta R.T. 0.000 min

Lab File: VN083708.D

Acq: 09 Sep 2024 10:03




Tgt Ion: 56 Resp: 57616


Ion Ratio Lower Upper

56 100

89 0.0 19.3 28.9#

84 88.6 71.3 106.9









#27
1,2-Dichloroethane-d4
Concen: 29.635 ug/l
RT: 8.577 min Scan# 1126
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 65 Resp: 84062
Ion Ratio Lower Upper
65 100
67 50.8 43.2 64.8



#28
1,4-Difluorobenzene
Concen: 30.000 ug/l
RT: 9.094 min Scan# 1214
Delta R.T. -0.006 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03

Tgt Ion: 114 Resp: 176340
Ion Ratio Lower Upper
114 100
63 21.7 15.9 23.9
88 16.4 12.7 19.1





#29

1,1-Dichloropropene

Concen: 17.204 ug/l

RT: 8.371 min Scan# 1090

Delta R.T. 0.006 min


Lab File: VN083708.D

Acq: 09 Sep 2024 10:03

Instrument :

MSVOA_N

ClientSampleId :






Tgt Ion: 75 Resp: 48120

Ion Ratio Lower Upper

75 100

110 32.7 0.0 70.2

77 30.5 0.0 61.4



#30

2-Butanone

Concen: 94.980 ug/l

RT: 7.483 min Scan# 940

Delta R.T. 0.000 min

Lab File: VN083708.D

Acq: 09 Sep 2024 10:03




Tgt Ion: 43 Resp: 125228


Ion Ratio Lower Upper

43 100

57 8.6 6.7 10.1

72 27.1 24.0 36.0









#31
2,2-Dichloropropane
Concen: 18.227 ug/l
RT: 7.489 min Scan# 941
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 77 Resp: 64997
Ion Ratio Lower Upper
77 100
97 21.5 18.3 27.5



#32
1,1,1-Trichloroethane
Concen: 17.775 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03


Tgt Ion: 97 Resp: 66397
Ion Ratio Lower Upper
97 100
99 65.4 51.0 76.4
61 44.0 33.8 50.8








#33
Carbon Tetrachloride
Concen: 17.937 ug/l
RT: 8.359 min Scan# 1109
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion:117 Resp: 58022
Ion Ratio Lower Upper
117 100
119 97.4 77.7 116.5
121 32.4 27.0 40.6



#34
Benzene
Concen: 18.012 ug/l
RT: 8.600 min Scan# 1130
Delta R.T. -0.006 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03


Tgt Ion: 78 Resp: 151159
Ion Ratio Lower Upper
78 100
77 25.1 19.2 28.8
51 20.0 13.3 19.9#








#35
Methacrylonitrile
Concen: 19.692 ug/l
RT: 7.777 min Scan# 990
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion:	41	Resp:	29187
Ion Ratio	Lower	Upper	
41	100		
39	56.3	25.3	75.9
67	68.2	38.7	116.1
52	29.8	13.6	40.8



#36
1,2-Dichloroethane
Concen: 18.100 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03

Tgt Ion:	62	Resp:	57105
Ion Ratio	Lower	Upper	
62	100		
98	8.5	8.2	12.4





#37

Trichloroethene

Concen: 17.789 ug/l

RT: 9.353 min Scan# 11

Delta R.T. 0.000 min

Lab File: VN083708.D

Acq: 09 Sep 2024 10:03

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion:130 Resp: 35056

Ion Ratio Lower Upper

130 100

95 94.2 80.2 120.4



#38

Methylcyclohexane

Concen: 17.705 ug/l

RT: 9.594 min Scan# 1299

Delta R.T. -0.006 min

Lab File: VN083708.D

Acq: 09 Sep 2024 10:03




Tgt Ion: 83 Resp: 59286


Ion Ratio Lower Upper

83 100

55 75.7 36.6 109.9

98 49.5 23.8 71.5





#39

1,2-Dichloropropane

Concen: 17.978 ug/l

RT: 9.618 min Scan# 11

Delta R.T. -0.006 min

Lab File: VN083708.D

Acq: 09 Sep 2024 10:03

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 63 Resp: 36311

Ion Ratio Lower Upper

63 100

65 32.8 25.0 37.4



#40

Dibromomethane

Concen: 18.018 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. -0.006 min

Lab File: VN083708.D

Acq: 09 Sep 2024 10:03




Tgt Ion: 93 Resp: 26610


Ion Ratio Lower Upper

93 100

95 80.1 0.0 164.6

174 87.7 0.0 193.2









#41
Bromodichloromethane
Concen: 17.723 ug/l
RT: 9.888 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 83 Resp: 55660
Ion Ratio Lower Upper
83 100
85 63.3 49.4 74.2
127 7.7 7.3 10.9



#42
Vinyl Acetate
Concen: 83.137 ug/l
RT: 6.600 min Scan# 790
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03


Tgt Ion: 43 Resp: 450676
Ion Ratio Lower Upper
43 100
86 10.6 9.4 14.2








#43
Ethyl Acetate
Concen: 18.379 ug/l
RT: 7.559 min Scan# 91
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 43 Resp: 48775
Ion Ratio Lower Upper
43 100
45 14.4 10.6 15.8



#44
Isopropyl Acetate
Concen: 18.730 ug/l
RT: 8.683 min Scan# 1144
Delta R.T. -0.006 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03


Tgt Ion: 43 Resp: 85971
Ion Ratio Lower Upper
43 100
61 25.7 25.0 37.4








#45
1,4-Dioxane
Concen: 420.192 ug/l
RT: 9.694 min Scan# 1316
Delta R.T. -0.006 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 88 Resp: 15043
Ion Ratio Lower Upper
88 100
43 26.6 19.8 29.6
58 70.1 55.4 83.0



#46
Methyl methacrylate
Concen: 19.109 ug/l
RT: 9.683 min Scan# 1314
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03

Tgt Ion: 41 Resp: 42453
Ion Ratio Lower Upper
41 100
69 78.5 51.1 153.3
39 60.2 25.8 77.3









#47
n-amyl acetate
Concen: 19.035 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 77395
Ion Ratio Lower Upper
43 100
55 23.2 23.4 35.2#



#48
t-1,3-Dichloropropene
Concen: 17.428 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03

Tgt Ion: 75 Resp: 57570
Ion Ratio Lower Upper
75 100
77 32.0 23.8 35.6









#49
 cis-1,3-Dichloropropene
 Concen: 18.011 ug/l
 RT: 10.312 min Scan# 1421
 Delta R.T. 0.000 min
 Lab File: VN083708.D
 Acq: 09 Sep 2024 10:03




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion: 75 Resp: 61875
 Ion Ratio Lower Upper
 75 100
 77 30.0 24.9 37.3
 39 51.9 31.0 46.6#



#50
 1,1,2-Trichloroethane
 Concen: 18.793 ug/l
 RT: 11.012 min Scan# 1540
 Delta R.T. -0.006 min
 Lab File: VN083708.D
 Acq: 09 Sep 2024 10:03

Tgt Ion: 97 Resp: 35316
 Ion Ratio Lower Upper
 97 100
 83 78.7 68.1 102.1
 85 51.2 41.8 62.6
 99 56.1 47.0 70.4









#51
Ethyl methacrylate
Concen: 18.117 ug/l
RT: 10.871 min Scan# 1516
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:	69	Resp:	58165
Ion Ratio	Lower	Upper	
69	100		
41	73.5	25.5	76.5
39	46.4	16.0	47.9



#52
1,3-Dichloropropane
Concen: 17.994 ug/l
RT: 11.159 min Scan# 1565
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03

Tgt Ion:	76	Resp:	61067
Ion Ratio	Lower	Upper	
76	100		
78	32.5	0.0	65.8





#53

Dibromochloromethane

Concen: 18.153 ug/l

RT: 11.359 min Scan# 1599

Delta R.T. 0.000 min

Lab File: VN083708.D

Acq: 09 Sep 2024 10:03

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion:129 Resp: 40918

Ion Ratio Lower Upper

129 100

127 75.5 39.4 118.1



#54

1,2-Dibromoethane

Concen: 18.113 ug/l

RT: 11.471 min Scan# 1618

Delta R.T. 0.006 min

Lab File: VN083708.D




Acq: 09 Sep 2024 10:03


Tgt Ion:107 Resp: 35213

Ion Ratio Lower Upper

107 100

109 92.0 74.7 112.1









#55
2-Chloroethyl vinyl ether
Concen: 85.564 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 63 Resp: 109480
Ion Ratio Lower Upper
63 100
65 32.8 26.2 39.2
106 28.4 23.8 35.8



#56
Bromoform
Concen: 18.061 ug/l
RT: 12.576 min Scan# 1806
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03

Tgt Ion:173 Resp: 25863
Ion Ratio Lower Upper
173 100
175 49.4 40.0 60.0
254 8.7 8.3 12.5





#57

Chlorobenzene-d5

Concen: 30.000 ug/l

RT: 11.865 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN083708.D

Acq: 09 Sep 2024 10:03

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:117 Resp: 157868

Ion Ratio Lower Upper

117 100

82 58.9 46.6 69.8

119 31.2 25.4 38.2



#58

4-Methyl-2-Pentanone

Concen: 99.295 ug/l

RT: 10.441 min Scan# 1443

Delta R.T. -0.006 min

Lab File: VN083708.D

Acq: 09 Sep 2024 10:03




Tgt Ion: 43 Resp: 248553


Ion Ratio Lower Upper

43 100

58 38.2 36.5 54.7


85 18.0 19.5 29.3#








#59
2-Hexanone
Concen: 100.397 ug/l
RT: 11.194 min Scan# 1157
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 43 Resp: 189895
Ion Ratio Lower Upper
43 100
58 52.3 50.4 75.6
57 19.3 18.6 27.8



#60
4-Bromofluorobenzene
Concen: 29.823 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03


Tgt Ion: 95 Resp: 81142
Ion Ratio Lower Upper
95 100
174 69.7 60.5 90.7
176 67.4 57.8 86.6








#61
Tetrachloroethene
Concen: 17.282 ug/l
RT: 11.100 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03


Instrument :
MSVOA_N
ClientSampleId :





Tgt Ion:164 Resp: 28874
Ion Ratio Lower Upper
164 100
166 129.9 106.7 160.1
129 95.4 85.8 128.6
131 97.8 79.1 118.7




#62
Toluene
Concen: 17.713 ug/l
RT: 10.630 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03



Tgt Ion: 91 Resp: 160730
Ion Ratio Lower Upper
91 100
92 57.9 46.5 69.7









#63
Toluene-d8
Concen: 30.093 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 98 Resp: 219979
Ion Ratio Lower Upper
98 100
100 64.6 52.7 79.1



#64
Chlorobenzene
Concen: 17.474 ug/l
RT: 11.888 min Scan# 1689
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03

Tgt Ion: 112 Resp: 98169
Ion Ratio Lower Upper
112 100
114 32.2 26.6 39.8









#65
 1,1,1,2-Tetrachloroethane
 Concen: 17.921 ug/l
 RT: 11.959 min Scan# 1701
 Delta R.T. 0.000 min
 Lab File: VN083708.D
 Acq: 09 Sep 2024 10:03




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion:	131	Resp:	35072
Ion Ratio	Lower	Upper	
131	100		
133	99.5	0.0	188.0
119	64.6	0.0	122.8



#66
 Ethyl Benzene
 Concen: 17.161 ug/l
 RT: 11.965 min Scan# 1702
 Delta R.T. 0.000 min
 Lab File: VN083708.D
 Acq: 09 Sep 2024 10:03

Tgt Ion:	91	Resp:	177087
Ion Ratio	Lower	Upper	
91	100		
106	28.1	24.4	36.6









#67
m/p-Xylenes
Concen: 34.537 ug/l
RT: 12.071 min Scan# 1720
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:106 Resp: 131117
Ion Ratio Lower Upper
106 100
91 214.5 166.6 249.8



#68
o-Xylene
Concen: 17.632 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03

Tgt Ion:106 Resp: 64783
Ion Ratio Lower Upper
106 100
91 228.5 110.0 330.0





#69

Styrene

Concen: 17.227 ug/l

RT: 12.412 min Scan# 1778

Delta R.T. 0.000 min


Lab File: VN083708.D

Acq: 09 Sep 2024 10:03

Instrument :

MSVOA_N

ClientSampleId :






Tgt Ion:104 Resp: 106951

Ion Ratio Lower Upper

104 100

78 55.1 42.6 64.0

103 54.9 43.7 65.5



#70

Isopropylbenzene

Concen: 17.654 ug/l

RT: 12.694 min Scan# 1826

Delta R.T. 0.000 min

Lab File: VN083708.D




Acq: 09 Sep 2024 10:03


Tgt Ion:105 Resp: 171180

Ion Ratio Lower Upper

105 100

120 25.2 17.8 33.1









#71
1,1,2,2-Tetrachloroethane
Concen: 19.099 ug/l
RT: 12.935 min Scan# 1867
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 83 Resp: 50581
Ion Ratio Lower Upper
83 100
131 9.3 5.2 15.6
85 63.7 32.3 96.8



#72
1,2,3-Trichloropropane
Concen: 29.158 ug/l
RT: 12.994 min Scan# 1877
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03

Tgt Ion: 75 Resp: 67704
Ion Ratio Lower Upper
75 100
77 132.0 0.0 367.6









#73
Bromobenzene
Concen: 17.709 ug/l
RT: 12.976 min Scan# 1874
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:156 Resp: 39441
Ion Ratio Lower Upper
156 100
77 226.7 0.0 437.6
158 93.5 0.0 203.8



#74
n-propylbenzene
Concen: 17.119 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03

Tgt Ion: 91 Resp: 194312
Ion Ratio Lower Upper
91 100
120 21.8 0.0 43.8









#75
2-Chlorotoluene
Concen: 17.673 ug/l
RT: 13.124 min Scan# 1899
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 91 Resp: 125200
Ion Ratio Lower Upper
91 100
126 30.0 0.0 61.2



#76
1,3,5-Trimethylbenzene
Concen: 17.665 ug/l
RT: 13.171 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03

Tgt Ion: 105 Resp: 143082
Ion Ratio Lower Upper
105 100
120 46.8 0.0 95.8









#77
t-1,4-Dichloro-2-butene
Concen: 17.263 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 75 Resp: 16943
Ion Ratio Lower Upper
75 100
53 111.1 45.9 137.7
89 52.8 25.9 77.7



#78
4-Chlorotoluene
Concen: 16.832 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03

Tgt Ion: 91 Resp: 120926
Ion Ratio Lower Upper
91 100
126 31.8 0.0 62.8









#79
tert-butylbenzene
Concen: 17.826 ug/l
RT: 13.435 min Scan# 1952
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:119 Resp: 126257
Ion Ratio Lower Upper
119 100
91 70.9 0.0 138.8
134 23.6 0.0 48.6



#80
1,2,4-Trimethylbenzene
Concen: 17.509 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03


Tgt Ion:105 Resp: 143312
Ion Ratio Lower Upper
105 100
120 44.1 0.0 86.8








#81
 sec-Butylbenzene
 Concen: 17.048 ug/l
 RT: 13.618 min Scan# 1983
 Delta R.T. 0.000 min
 Lab File: VN083708.D
 Acq: 09 Sep 2024 10:03


Instrument :
 MSVOA_N
 ClientSampleId :





Tgt Ion:105 Resp: 164949
 Ion Ratio Lower Upper
 105 100
 134 19.3 0.0 39.0




#82
 p-Isopropyltoluene
 Concen: 17.059 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. 0.000 min
 Lab File: VN083708.D
 Acq: 09 Sep 2024 10:03



Tgt Ion:119 Resp: 138870
 Ion Ratio Lower Upper
 119 100
 134 25.6 0.0 49.2
 91 26.3 0.0 48.8





#83

1,3-Dichlorobenzene

Concen: 16.865 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. 0.000 min

Lab File: VN083708.D

Acq: 09 Sep 2024 10:03

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:146 Resp: 72292

Ion Ratio Lower Upper

146 100

111 43.7 22.4 67.0

148 63.0 32.1 96.5



#84

1,4-Dichlorobenzene

Concen: 16.675 ug/l

RT: 13.812 min Scan# 2016

Delta R.T. 0.000 min

Lab File: VN083708.D

Acq: 09 Sep 2024 10:03




Tgt Ion:146 Resp: 71460


Ion Ratio Lower Upper

146 100

111 42.4 21.3 63.9

148 64.2 31.9 95.5





#85

n-Butylbenzene

Concen: 16.360 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. 0.000 min

Lab File: VN083708.D

Acq: 09 Sep 2024 10:03

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 91 Resp: 117524

Ion Ratio Lower Upper

91 100

92 53.0 0.0 105.0

134 23.8 0.0 51.8



#86

Hexachloroethane

Concen: 17.465 ug/l

RT: 14.335 min Scan# 2105

Delta R.T. 0.000 min

Lab File: VN083708.D




Acq: 09 Sep 2024 10:03


Tgt Ion: 117 Resp: 26943

Ion Ratio Lower Upper

117 100

201 65.9 38.0 114.0









#87
 1,2-Dichlorobenzene
 Concen: 17.203 ug/l
 RT: 14.106 min Scan# 2066
 Delta R.T. 0.000 min
 Lab File: VN083708.D
 Acq: 09 Sep 2024 10:03




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion:	146	Resp:	71564
Ion	Ratio	Lower	Upper
146	100		
111	44.5	22.6	67.8
148	63.8	31.3	93.8



#88
 1,2-Dibromo-3-Chloropropane
 Concen: 18.780 ug/l
 RT: 14.718 min Scan# 2170
 Delta R.T. 0.000 min
 Lab File: VN083708.D
 Acq: 09 Sep 2024 10:03


Tgt Ion:	75	Resp:	10930
Ion	Ratio	Lower	Upper
75	100		
155	71.9	61.7	92.5
157	97.1	83.6	125.4








#89
 1,2,4-Trichlorobenzene
 Concen: 15.888 ug/l
 RT: 15.841 min Scan# 21
 Delta R.T. 0.006 min
 Lab File: VN083708.D
 Acq: 09 Sep 2024 10:03


Instrument :
 MSVOA_N
 ClientSampleId :





Tgt Ion:180 Resp: 36209
 Ion Ratio Lower Upper
 180 100
 182 96.5 76.4 114.6
 145 37.0 28.4 42.6




#90
 Hexachlorobutadiene
 Concen: 17.038 ug/l
 RT: 15.500 min Scan# 2303
 Delta R.T. 0.000 min
 Lab File: VN083708.D
 Acq: 09 Sep 2024 10:03



Tgt Ion:225 Resp: 16072
 Ion Ratio Lower Upper
 225 100
 223 62.8 0.0 126.6
 227 63.9 0.0 126.6









#91
Naphthalene
Concen: 16.498 ug/l
RT: 15.641 min Scan# 21
Delta R.T. 0.000 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03

Instrument :
MSVOA_N
ClientSampleId :

Tgt Ion:128	Resp:	122921
Ion Ratio	Lower	Upper
128	100	
127	12.6	10.6 16.0
129	11.1	8.6 13.0



#92
1,2,3-Trichlorobenzene
Concen: 15.888 ug/l
RT: 15.841 min Scan# 2361
Delta R.T. 0.006 min
Lab File: VN083708.D
Acq: 09 Sep 2024 10:03

Tgt Ion:180	Resp:	36209
Ion Ratio	Lower	Upper
180	100	
182	96.5	0.0 191.0
145	37.0	0.0 71.0

