

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090924\
 Data File : VN083722.D
 Acq On : 09 Sep 2024 17:24
 Operator : JC\MD
 Sample : P3815-01DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OWBR-05-134.8-083024DL

Quant Time: Sep 10 02:13:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083024W.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 31 00:48:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	117993	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	239444	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	256268	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	118704	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	94976	57.594	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 115.180%			
35) Dibromofluoromethane	8.165	113	74306	51.560	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.120%			
50) Toluene-d8	10.565	98	265224	47.269	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 94.540%			
62) 4-Bromofluorobenzene	12.847	95	120245	57.838	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 115.680%			
Target Compounds						
12) 1,1-Dichloroethene	4.330	96	31649	24.484	ug/l	90
17) Carbon Disulfide	4.706	76	9528	2.529	ug/l #	92
21) trans-1,2-Dichloroethene	5.777	96	3175	2.358	ug/l	89
27) cis-1,2-Dichloroethene	7.482	96	28397	17.674	ug/l	89
42) 1,2-Dichloroethane	8.659	62	3315	1.322	ug/l	81
44) Trichloroethene	9.353	130	6827	4.197	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090924\
Data File : VN083722.D
Acq On : 09 Sep 2024 17:24
Operator : JC\MD
Sample : P3815-01DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OWBR-05-134.8-083024DL

Quant Time: Sep 10 02:13:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083024W.M
Quant Title : SW846 8260
QLast Update : Sat Aug 31 00:48:10 2024
Response via : Initial Calibration

