

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090924\
 Data File : VN083724.D
 Acq On : 09 Sep 2024 18:12
 Operator : JC\MD
 Sample : P3802-08RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OWBR-02-170-082924RE

Quant Time: Sep 10 02:14:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083024W.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 31 00:48:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	118460	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	244678	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	268058	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	130900	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	97598	58.951	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	117.900%
35) Dibromofluoromethane	8.165	113	79287	53.839	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.680%
50) Toluene-d8	10.565	98	274660	47.888	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.780%
62) 4-Bromofluorobenzene	12.847	95	130362	61.363	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	122.720%#
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	10226	15.489	ug/l	95
25) 2-Butanone	7.488	43	1789	1.990	ug/l	96
52) Toluene	10.623	92	3140	0.727	ug/l	97
69) o-Xylene	12.400	106	3430	0.878	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090924\
Data File : VN083724.D
Acq On : 09 Sep 2024 18:12
Operator : JC\MD
Sample : P3802-08RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OWBR-02-170-082924RE

Quant Time: Sep 10 02:14:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083024W.M
Quant Title : SW846 8260
QLast Update : Sat Aug 31 00:48:10 2024
Response via : Initial Calibration

