

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090924\
 Data File : VN083735.D
 Acq On : 09 Sep 2024 22:36
 Operator : JC\MD
 Sample : P3868-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OWBR-01-170-090524

Quant Time: Sep 10 02:19:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083024W.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 31 00:48:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	117617	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	238981	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	269062	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	124989	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	94932	57.751	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	115.500%
35) Dibromofluoromethane	8.165	113	76669	53.303	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.600%
50) Toluene-d8	10.565	98	270374	48.256	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.520%
62) 4-Bromofluorobenzene	12.847	95	124698	60.096	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	120.200%
Target Compounds						
					Qvalue	
27) cis-1,2-Dichloroethene	7.482	96	4287	2.677	ug/l	90
44) Trichloroethene	9.359	130	5027	3.096	ug/l	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090924\
Data File : VN083735.D
Acq On : 09 Sep 2024 22:36
Operator : JC\MD
Sample : P3868-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OWBR-01-170-090524

Quant Time: Sep 10 02:19:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083024W.M
Quant Title : SW846 8260
QLast Update : Sat Aug 31 00:48:10 2024
Response via : Initial Calibration

