

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090924\
 Data File : VN083710.D
 Acq On : 09 Sep 2024 10:50
 Operator : JC\MD
 Sample : VN0909WBL02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0909WBL02

Quant Time: Sep 10 01:17:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N090624W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Sep 06 23:42:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.812	128	29530	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.100	114	165572	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.865	117	139818	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.577	65	80558	29.844	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	99.467%
60) 4-Bromofluorobenzene	12.847	95	65858	27.330	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	91.100%
63) Toluene-d8	10.565	98	194929	30.109	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.367%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090924\
Data File : VN083710.D
Acq On : 09 Sep 2024 10:50
Operator : JC\MD
Sample : VN0909WBL02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0909WBL02

Quant Time: Sep 10 01:17:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N090624W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Sep 06 23:42:47 2024
Response via : Initial Calibration









#1
Bromochloromethane
Concen: 30.000 ug/l
RT: 7.812 min Scan# 996
Delta R.T. -0.000 min
Lab File: VN083710.D
Acq: 09 Sep 2024 10:50




Instrument :
MSVOA_N
ClientSampleId :
VN0909WBL02


Tgt Ion:128 Resp: 29530
Ion Ratio Lower Upper
128 100
49 170.0 0.0 403.7
130 125.3 0.0 323.3



#27
1,2-Dichloroethane-d4
Concen: 29.844 ug/l
RT: 8.577 min Scan# 1126
Delta R.T. 0.000 min
Lab File: VN083710.D
Acq: 09 Sep 2024 10:50

Tgt Ion: 65 Resp: 80558
Ion Ratio Lower Upper
65 100
67 49.6 43.2 64.8









#28
1,4-Difluorobenzene
Concen: 30.000 ug/l
RT: 9.100 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN083710.D
Acq: 09 Sep 2024 10:50




Instrument :
MSVOA_N
ClientSampleId :
VN0909WBL02


Tgt Ion:114 Resp: 165572
Ion Ratio Lower Upper
114 100
63 21.8 15.9 23.9
88 16.2 12.7 19.1



#57
Chlorobenzene-d5
Concen: 30.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. 0.000 min
Lab File: VN083710.D
Acq: 09 Sep 2024 10:50

Tgt Ion:117 Resp: 139818
Ion Ratio Lower Upper
117 100
82 59.1 46.6 69.8
119 32.1 25.4 38.2









#60
4-Bromofluorobenzene
Concen: 27.330 ug/l
RT: 12.847 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN083710.D
Acq: 09 Sep 2024 10:50

Instrument :
MSVOA_N
ClientSampleId :
VN0909WBL02

Tgt Ion: 95 Resp: 65858
Ion Ratio Lower Upper
95 100
174 71.6 60.5 90.7
176 67.4 57.8 86.6



#63
Toluene-d8
Concen: 30.109 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. 0.000 min
Lab File: VN083710.D
Acq: 09 Sep 2024 10:50

Tgt Ion: 98 Resp: 194929
Ion Ratio Lower Upper
98 100
100 66.2 52.7 79.1

