

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091018\
 Data File : VN051099.D
 Acq On : 11 Sep 2018 00:46
 Operator : MD\SY
 Sample : J4777-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PL-05-090718-A

Quant Time: Sep 11 05:27:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 11 02:27:09 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	627490	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	991316	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	846831	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	292654	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	380156	54.44	ug/l	0.00
Spiked Amount			Recovery	=	108.88%	
35) Dibromofluoromethane	7.59	113	369967	49.67	ug/l	0.00
Spiked Amount			Recovery	=	99.34%	
50) Toluene-d8	10.09	98	1353437	47.88	ug/l	0.00
Spiked Amount			Recovery	=	95.76%	
62) 4-Bromofluorobenzene	12.40	95	380127	40.65	ug/l	0.00
Spiked Amount			Recovery	=	81.30%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	13347	1.94	ug/l #	88
18) Methyl Acetate	4.34	43	8158	1.46	ug/l #	81
20) Methylene Chloride	4.55	84	217211	28.16	ug/l	98
43) Isopropyl Acetate	8.17	43	201007	18.07	ug/l #	85
44) Trichloroethene	8.83	130	10750	1.20	ug/l	98
95) Naphthalene	15.13	128	75941	11.50	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091018\
Data File : VN051099.D
Acq On : 11 Sep 2018 00:46
Operator : MD\SY
Sample : J4777-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PL-05-090718-A

Quant Time: Sep 11 05:27:15 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 11 02:27:09 2018
Response via : Initial Calibration

